

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ALAMANCE

Instrument Location Burlington PD


Instrument Serial No. 008907

267 W. Front St Burlington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008907

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 12:37pm        |
| AIR BLK         | .00        | 12:38pm        |
| ACCY CHK        | .08        | 12:38pm        |
| AIR BLK         | .00        | 12:40pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:40pm</b> |
| AIR BLK         | .00        | 12:41pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:42pm</b> |
| AIR BLK         | .00        | 12:43pm        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**ALAMANCE COUNTY BURLINGTON PD 000**

Serial Number: 008907      Test Record Number: 919  
Test Date: 02/27/2019      Test Time: 12:44pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:45pm |
| FLO  | Pass   | 12:45pm |
| FC   | Pass   | 12:45pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:45pm |
| SRC  | Pass   | 12:45pm |
| DET  | Pass   | 12:45pm |
| BAR  | Pass   | 12:45pm |
| BT   | Pass   | 12:45pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:45pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:45pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:46pm |
| CAL  | Pass   | 12:46pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County ALAMANCE

Instrument Location Burlington, PD


Instrument Serial No. 008812

267 W. Front St Burlington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Barnes  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008812

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:35pm |
| AIR BLK  | .00    | 12:36pm |
| ACCY CHK | .08    | 12:37pm |
| AIR BLK  | .00    | 12:38pm |
| SUB TEST | .00    | 12:39pm |
| AIR BLK  | .00    | 12:40pm |
| SUB TEST | .00    | 12:41pm |
| AIR BLK  | .00    | 12:42pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ALAMANCE COUNTY BURLINGTON PD 000

Serial Number: 008812      Test Record Number: 3365  
Test Date: 02/27/2019      Test Time: 12:43pm EST

System Check: *Passed*

Baseline Tests

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:44pm |
| FLO  | Pass   | 12:44pm |
| FC   | Pass   | 12:44pm |

Temperature Tests

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:44pm |
| SRC  | Pass   | 12:44pm |
| DET  | Pass   | 12:44pm |
| BAR  | Pass   | 12:44pm |
| BT   | Pass   | 12:44pm |

Blank Tests

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:44pm |


Printer Tests

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:44pm |

CRC Tests

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:45pm |
| CAL  | Pass   | 12:45pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ALAMANCE

Instrument Location ALAMANCE CO JAIL


Instrument Serial No. 008853

109 S. Maple St Graham NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008853

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 2:17pm        |
| AIR BLK         | .00        | 2:18pm        |
| ACCY CHK        | .07        | 2:18pm        |
| AIR BLK         | .00        | 2:20pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>2:20pm</b> |
| AIR BLK         | .00        | 2:21pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>2:22pm</b> |
| AIR BLK         | .00        | 2:23pm        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008853      Test Record Number: 2706  
Test Date: 02/27/2019      Test Time: 2:24pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:25pm |
| FLO  | Pass   | 2:25pm |
| FC   | Pass   | 2:25pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:25pm |
| SRC  | Pass   | 2:25pm |
| DET  | Pass   | 2:25pm |
| BAR  | Pass   | 2:25pm |
| BT   | Pass   | 2:25pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:25pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:25pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:26pm |
| CAL  | Pass   | 2:26pm |

Preventive Maintenance  
Status: *Pass*

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location ALAMANCE CO JAIL

Instrument Serial No. 008913 109 S. Maple St Graham, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902106


Exp Date: 01/21/2021

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:31pm |
| AIR BLK  | .00    | 2:31pm |
| ACCY CHK | .08    | 2:32pm |
| AIR BLK  | .00    | 2:33pm |
| SUB TEST | .00    | 2:34pm |
| AIR BLK  | .00    | 2:35pm |
| SUB TEST | .00    | 2:36pm |
| AIR BLK  | .00    | 2:37pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913      Test Record Number: 3429  
Test Date: 02/27/2019      Test Time: 2:38pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:38pm |
| FLO  | Pass   | 2:38pm |
| FC   | Pass   | 2:38pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:39pm |
| SRC  | Pass   | 2:39pm |
| DET  | Pass   | 2:39pm |
| BAR  | Pass   | 2:39pm |
| BT   | Pass   | 2:39pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:39pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:39pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:39pm |
| CAL  | Pass   | 2:39pm |

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Alexander Instrument Location Alexander County SO  
Instrument Serial No. 008813 91 Commercial Park Ave, Taylorsville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. J. Day  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALEXANDER COUNTY ALEXANDER COUNTY SD  
010

Serial Number: 008813

Test Date: 02/11/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 5:38pm |
| AIR BLK  | .00    | 5:39pm |
| ACCY CHK | .07    | 5:39pm |
| AIR BLK  | .00    | 5:41pm |
| SUB TEST | .00    | 5:41pm |
| AIR BLK  | .00    | 5:42pm |
| SUB TEST | .00    | 5:44pm |
| AIR BLK  | .00    | 5:44pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALEXANDER COUNTY ALEXANDER COUNTY SD 010

Serial Number: 008813

Test Record Number: 1833

Test Date: 02/11/2019

Test Time: 5:45pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 5:45pm |
| FLO  | Pass   | 5:45pm |
| FC   | Pass   | 5:45pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 5:45pm |
| SRC  | Pass   | 5:45pm |
| DET  | Pass   | 5:45pm |
| BAR  | Pass   | 5:45pm |
| BT   | Pass   | 5:45pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 5:46pm |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 5:46pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 5:46pm |
| CAL  | Pass   | 5:46pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Avery

Instrument Location Avery Co. Jail


Instrument Serial No. 008664

Newland NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:21am |
| AIR BLK  | .00    | 11:22am |
| ACCY CHK | .08    | 11:23am |
| AIR BLK  | .00    | 11:24am |
| SUB TEST | .00    | 11:25am |
| AIR BLK  | .00    | 11:25am |
| SUB TEST | .00    | 11:27am |
| AIR BLK  | .00    | 11:28am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664      Test Record Number: 963  
Test Date: 02/22/2019      Test Time: 11:29am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:29am |
| FLO  | Pass   | 11:29am |
| FC   | Pass   | 11:29am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:29am |
| SRC  | Pass   | 11:29am |
| DET  | Pass   | 11:29am |
| BAR  | Pass   | 11:29am |
| BT   | Pass   | 11:29am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:30am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:30am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:30am |
| CAL  | Pass   | 11:30am |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Beaufort Instrument Location Belhaven PD  
Instrument Serial No. 008928 Belhaven, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>nd</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key M O  
Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: 008928

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:02am |
| AIR BLK  | .00    | 11:03am |
| ACCY CHK | .08    | 11:03am |
| AIR BLK  | .00    | 11:04am |
| SUB TEST | .00    | 11:05am |
| AIR BLK  | .00    | 11:06am |
| SUB TEST | .00    | 11:07am |
| AIR BLK  | .00    | 11:08am |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: 008928      Test Record Number: 378  
Test Date: 02/22/2019      Test Time: 11:17am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:18am |
| FLO  | Pass   | 11:18am |
| FC   | Pass   | 11:18am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:18am |
| SRC  | Pass   | 11:18am |
| DET  | Pass   | 11:18am |
| BAR  | Pass   | 11:18am |
| BT   | Pass   | 11:18am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:19am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:19am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:19am |
| CAL  | Pass   | 11:19am |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Cabarrus Instrument Location Cabarrus County/50  
Instrument Serial No. 008792 30 Corban Ave, Concord

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CABARRUS COUNTY CABARRUS COUNTY SO 120

Serial Number: 008792

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG722408

Exp Date: 08/12/2019

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 11:30am        |
| AIR BLK         | .00        | 11:30am        |
| ACCY CHK        | .08        | 11:31am        |
| AIR BLK         | .00        | 11:32am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:33am</b> |
| AIR BLK         | .00        | 11:34am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:35am</b> |
| AIR BLK         | .00        | 11:36am        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**Intox EC/IR-II: Preventive Maintenance**

CABARRUS COUNTY CABARRUS COUNTY SO 120

Serial Number: 008792      Test Record Number: 2957  
Test Date: 02/15/2019      Test Time: 11:37am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:37am |
| FLO  | Pass   | 11:37am |
| FC   | Pass   | 11:38am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:38am |
| SRC  | Pass   | 11:38am |
| DET  | Pass   | 11:38am |
| BAR  | Pass   | 11:38am |
| BT   | Pass   | 11:38am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:38am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:38am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:39am |
| CAL  | Pass   | 11:39am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cabarrus Instrument Location Cabarrus County SO  
Instrument Serial No. 008625 30 Corban Ave., Concord

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. D. Hays  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CABARRUS COUNTY CABARRUS COUNTY SO 120

Serial Number: 008625

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805802


Exp Date: 02/27/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:43am |
| AIR BLK  | .00    | 11:44am |
| ACCY CHK | .08    | 11:45am |
| AIR BLK  | .00    | 11:46am |
| SUB TEST | .00    | 11:47am |
| AIR BLK  | .00    | 11:48am |
| SUB TEST | .00    | 11:49am |
| AIR BLK  | .00    | 11:50am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CABARRUS COUNTY CABARRUS COUNTY SO 120

Serial Number: 008625      Test Record Number: 5122  
Test Date: 02/15/2019      Test Time: 11:51am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:52am |
| FLO  | Pass   | 11:52am |
| FC   | Pass   | 11:52am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:52am |
| SRC  | Pass   | 11:52am |
| DET  | Pass   | 11:52am |
| BAR  | Pass   | 11:52am |
| BT   | Pass   | 11:52am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:53am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:53am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:53am |
| CAL  | Pass   | 11:53am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Camden Instrument Location Camden Co. S.O.  
Instrument Serial No. 008940 113 Hwy 343, Camden, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940

Test Date: 02/13/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:14am |
| AIR BLK  | .00    | 11:15am |
| ACCY CHK | .08    | 11:15am |
| AIR BLK  | .00    | 11:16am |
| SUB TEST | .00    | 11:17am |
| AIR BLK  | .00    | 11:18am |
| SUB TEST | .00    | 11:19am |
| AIR BLK  | .00    | 11:20am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940      Test Record Number: 928  
Test Date: 02/13/2019      Test Time: 11:21am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:22am |
| FLO  | Pass   | 11:22am |
| FC   | Pass   | 11:22am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:22am |
| SRC  | Pass   | 11:22am |
| DET  | Pass   | 11:22am |
| BAR  | Pass   | 11:22am |
| BT   | Pass   | 11:22am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:22am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:22am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:23am |
| CAL  | Pass   | 11:23am |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Chatham Co. Instrument Location Siler City Police Dept.  
Instrument Serial No. 008811 Siler City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CHATHAM COUNTY SILER CITY PD 180

Serial Number: 008811

Test Date: 02/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG807102


Exp Date: 03/12/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:03pm |
| AIR BLK  | .00    | 12:04pm |
| ACCY CHK | .08    | 12:05pm |
| AIR BLK  | .00    | 12:05pm |
| SUB TEST | .00    | 12:06pm |
| AIR BLK  | .00    | 12:07pm |
| SUB TEST | .00    | 12:09pm |
| AIR BLK  | .00    | 12:10pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**CHATHAM COUNTY SILER CITY PD 180**

Serial Number: 008811      Test Record Number: 1337

Test Date: 02/25/2019      Test Time: 12:12pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:12pm |
| FLO  | Pass   | 12:12pm |
| FC   | Pass   | 12:12pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:13pm |
| SRC  | Pass   | 12:13pm |
| DET  | Pass   | 12:13pm |
| BAR  | Pass   | 12:13pm |
| BT   | Pass   | 12:13pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:13pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:13pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:13pm |
| CAL  | Pass   | 12:13pm |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chatham Co. Instrument Location Chatham Co. Detention Center

Instrument Serial No. 008591 P. Hsboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28<sup>th</sup> day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CHATHAM COUNTY DETENTION CENTER 180

Serial Number: 008591  
Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:56pm |
| AIR BLK  | .00    | 1:57pm |
| ACCY CHK | .08    | 1:57pm |
| AIR BLK  | .00    | 1:58pm |
| SUB TEST | .00    | 1:59pm |
| AIR BLK  | .00    | 2:00pm |
| SUB TEST | .00    | 2:01pm |
| AIR BLK  | .00    | 2:02pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**CHATHAM COUNTY DETENTION CENTER 180**

Serial Number: 008591      Test Record Number: 2121  
Test Date: 02/28/2019      Test Time: 2:03pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:03pm |
| FLO  | Pass   | 2:03pm |
| FC   | Pass   | 2:03pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:03pm |
| SRC  | Pass   | 2:03pm |
| DET  | Pass   | 2:03pm |
| BAR  | Pass   | 2:03pm |
| BT   | Pass   | 2:03pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:04pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:04pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:04pm |
| CAL  | Pass   | 2:04pm |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cleveland Instrument Location Kings Mountain PD  
Instrument Serial No. 008900 112 S. Piedmont Ave., Kings Mountain

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**CLEVELAND COUNTY KINGS MOUNTAIN PD 220**

Serial Number: 008900

Test Date: 02/05/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:35am |
| AIR BLK  | .00    | 9:36am |
| ACCY CHK | .08    | 9:37am |
| AIR BLK  | .00    | 9:38am |
| SUB TEST | .00    | 9:39am |
| AIR BLK  | .00    | 9:40am |
| SUB TEST | .00    | 9:41am |
| AIR BLK  | .00    | 9:42am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY KINGS MOUNTAIN PD 220

Serial Number: 008900      Test Record Number: 750  
Test Date: 02/05/2019      Test Time: 9:45am EST

System Check: Passed

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:45am |
| FLO  | Pass   | 9:45am |
| FC   | Pass   | 9:45am |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:45am |
| SRC  | Pass   | 9:45am |
| DET  | Pass   | 9:45am |
| BAR  | Pass   | 9:45am |
| BT   | Pass   | 9:45am |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:46am |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:46am |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:46am |
| CAL  | Pass   | 9:46am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Cleveland Instrument Location Cleveland County SO-Annex  
Instrument Serial No. 008887 407 M'Brayer St., Shelby

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND SO-ANNEX  
220

Serial Number: 008887  
Test Date: 02/25/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801


Exp Date: 02/27/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:19am |
| AIR BLK  | .00    | 9:20am |
| ACCY CHK | .08    | 9:21am |
| AIR BLK  | .00    | 9:22am |
| SUB TEST | .00    | 9:22am |
| AIR BLK  | .00    | 9:23am |
| SUB TEST | .00    | 9:25am |
| AIR BLK  | .00    | 9:26am |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

CLEVELAND COUNTY CLEVELAND SO-ANNEX 220

Serial Number: 008887      Test Record Number: 2850  
Test Date: 02/25/2019      Test Time: 9:27am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:27am |
| FLO  | Pass   | 9:27am |
| FC   | Pass   | 9:27am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:27am |
| SRC  | Pass   | 9:27am |
| DET  | Pass   | 9:27am |
| BAR  | Pass   | 9:27am |
| BT   | Pass   | 9:27am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:28am |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:28am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:28am |
| CAL  | Pass   | 9:28am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location CUMBERLAND CO  
Instrument Serial No. 008614 DETENTION CENTER  
FAYETTEVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Rg Banno

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614  
Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202


Exp Date: 01/22/2021

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:45am |
| AIR BLK  | .00    | 9:45am |
| ACCY CHK | .07    | 9:46am |
| AIR BLK  | .00    | 9:47am |
| SUB TEST | .00    | 9:48am |
| AIR BLK  | .00    | 9:49am |
| SUB TEST | .00    | 9:50am |
| AIR BLK  | .00    | 9:51am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

**CUMBERLAND COUNTY DETENTION CENTER 250**

Serial Number: 008614      Test Record Number: 4247  
Test Date: 02/18/2019      Test Time: 9:52am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:52am |
| FLO  | Pass   | 9:52am |
| FC   | Pass   | 9:52am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:52am |
| SRC  | Pass   | 9:52am |
| DET  | Pass   | 9:52am |
| BAR  | Pass   | 9:52am |
| BT   | Pass   | 9:52am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:53am |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:53am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:53am |
| CAL  | Pass   | 9:53am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location CUMBERLAND CO  
Instrument Serial No. 008721 DETENTION CENTER  
FAYETTEVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008721

Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202


Exp Date: 01/22/2021

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:45am |
| AIR BLK  | .00    | 10:46am |
| ACCY CHK | .07    | 10:46am |
| AIR BLK  | .00    | 10:47am |
| SUB TEST | .00    | 10:48am |
| AIR BLK  | .00    | 10:49am |
| SUB TEST | .00    | 10:50am |
| AIR BLK  | .00    | 10:51am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008721      Test Record Number: 1042

Test Date: 02/18/2019      Test Time: 10:52am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:53am |
| FLO  | Pass   | 10:53am |
| FC   | Pass   | 10:53am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:53am |
| SRC  | Pass   | 10:53am |
| DET  | Pass   | 10:53am |
| BAR  | Pass   | 10:53am |
| BT   | Pass   | 10:53am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:53am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:53am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:54am |
| CAL  | Pass   | 10:54am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CUMBERLAND Instrument Location CUMBERLAND CO

Instrument Serial No. 008672 DETENTION CENTER  
FAYETTEVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bane  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**CUMBERLAND COUNTY DETENTION CENTER 250**

Serial Number: 008672

Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202


Exp Date: 01/22/2021

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:22am |
| AIR BLK  | .00    | 11:22am |
| ACCY CHK | .07    | 11:23am |
| AIR BLK  | .00    | 11:24am |
| SUB TEST | .00    | 11:25am |
| AIR BLK  | .00    | 11:25am |
| SUB TEST | .00    | 11:27am |
| AIR BLK  | .00    | 11:28am |

- Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672      Test Record Number: 6408  
Test Date: 02/18/2019      Test Time: 11:30am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:30am |
| FLO  | Pass   | 11:30am |
| FC   | Pass   | 11:30am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:30am |
| SRC  | Pass   | 11:30am |
| DET  | Pass   | 11:30am |
| BAR  | Pass   | 11:30am |
| BT   | Pass   | 11:30am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:31am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:31am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:31am |
| CAL  | Pass   | 11:31am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND

Instrument Location CUMBERLAND CO


Instrument Serial No. 008787

DETENTION CENTER  
FAYETTEVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008787

Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202


Exp Date: 01/22/2021

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 12:11pm        |
| AIR BLK         | .00        | 12:12pm        |
| ACCY CHK        | .08        | 12:13pm        |
| AIR BLK         | .00        | 12:13pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:14pm</b> |
| AIR BLK         | .00        | 12:15pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:16pm</b> |
| AIR BLK         | .00        | 12:17pm        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008787      Test Record Number: 841  
Test Date: 02/18/2019      Test Time: 12:18pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:18pm |
| FLO  | Pass   | 12:18pm |
| FC   | Pass   | 12:18pm |

Temperature Tests

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:18pm |
| SRC  | Pass   | 12:18pm |
| DET  | Pass   | 12:18pm |
| BAR  | Pass   | 12:18pm |
| BT   | Pass   | 12:18pm |

Blank Tests

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:19pm |


Printer Tests

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:19pm |

CRC Tests

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:19pm |
| CAL  | Pass   | 12:19pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location FT BRAGG  
Instrument Serial No. 008908 PMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Rg Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY FT BRAGG LEC 250

Serial Number: 008908

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:00pm |
| AIR BLK  | .00    | 2:01pm |
| ACCY CHK | .08    | 2:01pm |
| AIR BLK  | .00    | 2:03pm |
| SUB TEST | .00    | 2:03pm |
| AIR BLK  | .00    | 2:04pm |
| SUB TEST | .00    | 2:06pm |
| AIR BLK  | .00    | 2:06pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY FT BRAGG LEC 250

Serial Number: 008908      Test Record Number: 1727  
Test Date: 02/21/2019      Test Time: 2:07pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:08pm |
| FLO  | Pass   | 2:08pm |
| FC   | Pass   | 2:08pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:08pm |
| SRC  | Pass   | 2:08pm |
| DET  | Pass   | 2:08pm |
| BAR  | Pass   | 2:08pm |
| BT   | Pass   | 2:08pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:08pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:08pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:09pm |
| CAL  | Pass   | 2:09pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location FT BRAGG  
Instrument Serial No. 008903 PMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 008903

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 2:17pm        |
| AIR BLK         | .00        | 2:18pm        |
| ACCY CHK        | .07        | 2:18pm        |
| AIR BLK         | .00        | 2:19pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>2:20pm</b> |
| AIR BLK         | .00        | 2:21pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>2:22pm</b> |
| AIR BLK         | .00        | 2:23pm        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 008903      Test Record Number: 2331  
Test Date: 02/21/2019      Test Time: 2:24pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:24pm |
| FLO  | Pass   | 2:24pm |
| FC   | Pass   | 2:24pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:24pm |
| SRC  | Pass   | 2:24pm |
| DET  | Pass   | 2:24pm |
| BAR  | Pass   | 2:24pm |
| BT   | Pass   | 2:24pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:25pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:25pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:25pm |
| CAL  | Pass   | 2:25pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Currituck Instrument Location Currituck Co. S.O.  
Instrument Serial No. DD8947 407-A Maple Rd., Maple, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**CURRITUCK COUNTY CURRITUCK SO-MAPLE**  
**260**

Serial Number: 008947  
Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

**PREVENTIVE, MAINTENANCE**

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:16am |
| AIR BLK  | .00    | 10:17am |
| ACCY CHK | .08    | 10:18am |
| AIR BLK  | .00    | 10:19am |
| SUB TEST | .00    | 10:20am |
| AIR BLK  | .00    | 10:20am |
| SUB TEST | .00    | 10:22am |
| AIR BLK  | .00    | 10:23am |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**CURRITUCK COUNTY CURRITUCK SO-MAPLE 260**

Serial Number: 008947      Test Record Number: 2338  
Test Date: 02/26/2019      Test Time: 10:24am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:24am |
| FLO  | Pass   | 10:24am |
| FC   | Pass   | 10:24am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:25am |
| SRC  | Pass   | 10:25am |
| DET  | Pass   | 10:25am |
| BAR  | Pass   | 10:25am |
| BT   | Pass   | 10:25am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:25am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:25am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:25am |
| CAL  | Pass   | 10:25am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Dare Instrument Location Dare Co Detention Ctr.  
Instrument Serial No. 008783 1044 Driftwood Dr., Manteo, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: 008783  
Test Date: 02/18/2019

Citation Number: M00000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G  
Permit Number: 12955E  
Effective:  
06/01/2017-06/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG734101  
Exp Date: 12/07/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:02am |
| AIR BLK  | .00    | 11:03am |
| ACCY CHK | .07    | 11:03am |
| AIR BLK  | .00    | 11:05am |
| SUB TEST | .00    | 11:05am |
| AIR BLK  | .00    | 11:06am |
| SUB TEST | .00    | 11:08am |
| AIR BLK  | .00    | 11:08am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**DARE COUNTY DARE CO DETENTION CE 270**

Serial Number: 008783      Test Record Number: 814  
Test Date: 02/18/2019      Test Time: 11:09am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:10am |
| FLO  | Pass   | 11:10am |
| FC   | Pass   | 11:10am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:10am |
| SRC  | Pass   | 11:10am |
| DET  | Pass   | 11:10am |
| BAR  | Pass   | 11:10am |
| BT   | Pass   | 11:10am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:11am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:11am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:11am |
| CAL  | Pass   | 11:11am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Dare Instrument Location Dare Co. Detention Ctr.  
Instrument Serial No. 008804 1044 Driftwood Dr., Manteo, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key R  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: 008804  
Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902106

Exp Date: 01/21/2021

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 11:13am        |
| AIR BLK         | .00        | 11:14am        |
| ACCY CHK        | .08        | 11:15am        |
| AIR BLK         | .00        | 11:16am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:16am</b> |
| AIR BLK         | .00        | 11:17am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:19am</b> |
| AIR BLK         | .00        | 11:19am        |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*DARE COUNTY DARE CO DETENTION CE 270*

Serial Number: 008804      Test Record Number: 2174  
Test Date: 02/18/2019      Test Time: 11:20am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:21am |
| FLO  | Pass   | 11:21am |
| FC   | Pass   | 11:21am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:21am |
| SRC  | Pass   | 11:21am |
| DET  | Pass   | 11:21am |
| BAR  | Pass   | 11:21am |
| BT   | Pass   | 11:21am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:22am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:22am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:22am |
| CAL  | Pass   | 11:22am |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County DARE Instrument Location Dare C.O.S.D. - Hatteras  
Instrument Serial No. 008807 50346 NC Hwy 12, Frisco, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>nd</sup> day of FEBRUARY, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda Kuehl  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DARE COUNTY DARE CO SO HATTERAS 270

Serial Number: 008807

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:28am |
| AIR BLK  | .00    | 10:28am |
| ACCY CHK | .07    | 10:29am |
| AIR BLK  | .00    | 10:30am |
| SUB TEST | .00    | 10:31am |
| AIR BLK  | .00    | 10:32am |
| SUB TEST | .00    | 10:33am |
| AIR BLK  | .00    | 10:34am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

DARE COUNTY DARE CO SO HATTERAS 270

Serial Number: 008807      Test Record Number: 1018  
Test Date: 02/22/2019      Test Time: 10:35am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:35am |
| FLO  | Pass   | 10:35am |
| FC   | Pass   | 10:35am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:35am |
| SRC  | Pass   | 10:35am |
| DET  | Pass   | 10:35am |
| BAR  | Pass   | 10:35am |
| BT   | Pass   | 10:35am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:36am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:36am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:36am |
| CAL  | Pass   | 10:36am |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location Lexington  
Instrument Serial No. 008883 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008883

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:44pm |
| AIR BLK  | .00    | 2:45pm |
| ACCY CHK | .08    | 2:45pm |
| AIR BLK  | .00    | 2:46pm |
| SUB TEST | .00    | 2:47pm |
| AIR BLK  | .00    | 2:48pm |
| SUB TEST | .00    | 2:49pm |
| AIR BLK  | .00    | 2:50pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008883      Test Record Number: 2043

Test Date: 02/15/2019      Test Time: 2:51pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:51pm |
| FLO  | Pass   | 2:51pm |
| FC   | Pass   | 2:51pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:51pm |
| SRC  | Pass   | 2:51pm |
| DET  | Pass   | 2:51pm |
| BAR  | Pass   | 2:51pm |
| BT   | Pass   | 2:51pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:52pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:52pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:53pm |
| CAL  | Pass   | 2:53pm |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location THOMASVILLE  
Instrument Serial No. 008872 POLICE DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: 008872

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:51pm |
| AIR BLK  | .00    | 1:52pm |
| ACCY CHK | .07    | 1:52pm |
| AIR BLK  | .00    | 1:53pm |
| SUB TEST | .00    | 1:54pm |
| AIR BLK  | .00    | 1:55pm |
| SUB TEST | .00    | 1:57pm |
| AIR BLK  | .00    | 1:58pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: 008872      Test Record Number: 1423  
Test Date: 02/15/2019      Test Time: 1:59pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:59pm |
| FLO  | Pass   | 1:59pm |
| FC   | Pass   | 1:59pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:59pm |
| SRC  | Pass   | 1:59pm |
| DET  | Pass   | 1:59pm |
| BAR  | Pass   | 1:59pm |
| BT   | Pass   | 1:59pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:00pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:00pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:00pm |
| CAL  | Pass   | 2:00pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Davie Instrument Location Davie County Jail  
Instrument Serial No. 008905 Mocksville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26<sup>th</sup> day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

\_\_\_\_\_  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905

Test Date: 02/26/2019

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821801


Exp Date: 08/06/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:16pm |
| AIR BLK  | .00    | 1:17pm |
| ACCY CHK | .08    | 1:17pm |
| AIR BLK  | .00    | 1:19pm |
| SUB TEST | .00    | 1:19pm |
| AIR BLK  | .00    | 1:20pm |
| SUB TEST | .00    | 1:22pm |
| AIR BLK  | .00    | 1:23pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905      Test Record Number: 2326  
Test Date: 02/26/2019      Test Time: 1:24pm EST

System Check: *Passed*

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:24pm |
| FLO  | Pass   | 1:24pm |
| FC   | Pass   | 1:24pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:24pm |
| SRC  | Pass   | 1:24pm |
| DET  | Pass   | 1:24pm |
| BAR  | Pass   | 1:24pm |
| BT   | Pass   | 1:24pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:25pm |

Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:25pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:25pm |
| CAL  | Pass   | 1:25pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Wallace  
Instrument Serial No. 008858 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858

Test Date: 02/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:41pm |
| AIR BLK  | .00    | 12:42pm |
| ACCY CHK | .08    | 12:42pm |
| AIR BLK  | .00    | 12:43pm |
| SUB TEST | .00    | 12:44pm |
| AIR BLK  | .00    | 12:45pm |
| SUB TEST | .00    | 12:46pm |
| AIR BLK  | .00    | 12:47pm |

Reported AC: 00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**DUPLIN COUNTY WALLACE PD 300**

Serial Number: 008858      Test Record Number: 892  
Test Date: 02/01/2019      Test Time: 12:48pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:49pm |
| FLO  | Pass   | 12:49pm |
| FC   | Pass   | 12:49pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:49pm |
| SRC  | Pass   | 12:49pm |
| DET  | Pass   | 12:49pm |
| BAR  | Pass   | 12:49pm |
| BT   | Pass   | 12:49pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:50pm |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:50pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:50pm |
| CAL  | Pass   | 12:50pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Duplin County  
Instrument Serial No. 008864 Sheriffs Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes

Signature of Certifying Official

601

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY DUPLIN CO SD 300

Serial Number: 008864

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:00am |
| AIR BLK  | .00    | 11:01am |
| ACCY CHK | .08    | 11:02am |
| AIR BLK  | .00    | 11:03am |
| SUB TEST | .00    | 11:04am |
| AIR BLK  | .00    | 11:04am |
| SUB TEST | .00    | 11:06am |
| AIR BLK  | .00    | 11:07am |

Reported AC: 00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*DUPLIN COUNTY DUPLIN CO SD 300*

Serial Number: 008864      Test Record Number: 3571  
Test Date: 02/21/2019      Test Time: 11:07am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:08am |
| FLO  | Pass   | 11:08am |
| FC   | Pass   | 11:08am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:08am |
| SRC  | Pass   | 11:08am |
| DET  | Pass   | 11:08am |
| BAR  | Pass   | 11:08am |
| BT   | Pass   | 11:08am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:09am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:09am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:09am |
| CAL  | Pass   | 11:09am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County

Forsyth

Instrument Location

Forsyth County

Instrument Serial No.

008925

Detention  
Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Wale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**FORSYTH COUNTY FORSYTH CO DETENTION**  
330

Serial Number: 008925  
Test Date: 02/15/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: FARLEY, CYNTHIA D  
Permit Number: 24123E  
Effective:  
11/01/2018-11/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG831801  
Exp Date: 11/14/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:39pm |
| AIR BLK  | .00    | 1:40pm |
| ACCY CHK | .08    | 1:40pm |
| AIR BLK  | .00    | 1:41pm |
| SUB TEST | .00    | 1:43pm |
| AIR BLK  | .00    | 1:44pm |
| SUB TEST | .00    | 1:45pm |
| AIR BLK  | .00    | 1:46pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008925      Test Record Number: 2587  
Test Date: 02/15/2019      Test Time: 1:47pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:48pm |
| FLO  | Pass   | 1:48pm |
| FC   | Pass   | 1:48pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:48pm |
| SRC  | Pass   | 1:48pm |
| DET  | Pass   | 1:48pm |
| BAR  | Pass   | 1:48pm |
| BT   | Pass   | 1:48pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:49pm |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:49pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:49pm |
| CAL  | Pass   | 1:49pm |

Preventive Maintenance  
Status: Pass

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County  
Instrument Serial No. 008583 Detention  
Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



C. Dale Farker  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION  
330

Serial Number: 008583  
Test Date: 02/15/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: FARLEY, CYNTHIA D  
Permit Number: 24123E  
Effective:  
11/01/2018-11/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG831801  
Exp Date: 11/14/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:44pm |
| AIR BLK  | .00    | 1:45pm |
| ACCY CHK | .08    | 1:45pm |
| AIR BLK  | .00    | 1:47pm |
| SUB TEST | .00    | 1:48pm |
| AIR BLK  | .00    | 1:49pm |
| SUB TEST | .00    | 1:51pm |
| AIR BLK  | .00    | 1:52pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**FORSYTH COUNTY FORSYTH CO DETENTION 330**

Serial Number: 008583      Test Record Number: 7531  
Test Date: 02/15/2019      Test Time: 1:53pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:53pm |
| FLO  | Pass   | 1:53pm |
| FC   | Pass   | 1:54pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:54pm |
| SRC  | Pass   | 1:54pm |
| DET  | Pass   | 1:54pm |
| BAR  | Pass   | 1:54pm |
| BT   | Pass   | 1:54pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:54pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:54pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:55pm |
| CAL  | Pass   | 1:55pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

**This form is used when performing Preventive Maintenance procedures**  
**Forensic Tests for Alcohol Branch**  
**Department of Health and Human Services**  
**Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County  
Instrument Serial No. 008659 Detention  
Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**FORSYTH COUNTY FORSYTH CO DETENTION**  
330

Serial Number: 008659  
Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG831801

Exp Date: 11/14/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:37pm |
| AIR BLK  | .00    | 1:38pm |
| ACCY CHK | .08    | 1:38pm |
| AIR BLK  | .00    | 1:40pm |
| SUB TEST | .00    | 1:40pm |
| AIR BLK  | .00    | 1:41pm |
| SUB TEST | .00    | 1:45pm |
| AIR BLK  | .00    | 1:46pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**  
**Forensic Tests for Alcohol Branch**  
**Department of Health and Human Services**  
**Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008659      Test Record Number: 5086  
Test Date: 02/15/2019      Test Time: 1:47pm EST

System Check: *Passed*

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:47pm |
| FLO  | Pass   | 1:47pm |
| FC   | Pass   | 1:47pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:47pm |
| SRC  | Pass   | 1:47pm |
| DET  | Pass   | 1:47pm |
| BAR  | Pass   | 1:47pm |
| BT   | Pass   | 1:47pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:48pm |

Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:48pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:48pm |
| CAL  | Pass   | 1:48pm |

Preventive Maintenance  
Status: Pass

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Kernersville  
Instrument Serial No. 008650 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27<sup>th</sup> day of February 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



C. Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:35pm |
| AIR BLK  | .00    | 1:36pm |
| ACCY CHK | .07    | 1:36pm |
| AIR BLK  | .00    | 1:37pm |
| SUB TEST | .00    | 1:39pm |
| AIR BLK  | .00    | 1:40pm |
| SUB TEST | .00    | 1:42pm |
| AIR BLK  | .00    | 1:43pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*FORSYTH COUNTY KERNERSVILLE PD 330*

Serial Number: 008650      Test Record Number: 1517  
Test Date: 02/27/2019      Test Time: 1:43pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:44pm |
| FLO  | Pass   | 1:44pm |
| FC   | Pass   | 1:44pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:44pm |
| SRC  | Pass   | 1:44pm |
| DET  | Pass   | 1:44pm |
| BAR  | Pass   | 1:44pm |
| BT   | Pass   | 1:44pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:45pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:45pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:45pm |
| CAL  | Pass   | 1:45pm |

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures**  
**Forensic Tests for Alcohol Branch**  
**Department of Health and Human Services**  
**Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location But Mobile Unit 2

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris Dwyer

Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**FORSYTH BAT MOBILE UNIT 2 330**

Serial Number: 008871

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 9:04pm        |
| AIR BLK         | .00        | 9:05pm        |
| ACCY CHK        | .08        | 9:06pm        |
| AIR BLK         | .00        | 9:07pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:07pm</b> |
| AIR BLK         | .00        | 9:08pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:10pm</b> |
| AIR BLK         | .00        | 9:10pm        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**FORSYTH BAT MOBILE UNIT 2 330**

Serial Number: 008871      Test Record Number: 952  
Test Date: 02/27/2019      Test Time: 9:11pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:12pm |
| FLO  | Pass   | 9:12pm |
| FC   | Pass   | 9:12pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:12pm |
| SRC  | Pass   | 9:12pm |
| DET  | Pass   | 9:12pm |
| BAR  | Pass   | 9:12pm |
| BT   | Pass   | 9:12pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:13pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:13pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:13pm |
| CAL  | Pass   | 9:13pm |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County Gaston Instrument Location Blat Mobile Unit 2

Instrument Serial No. 008871

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chloe Joy  
Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GASTON BAT MOBILE UNIT 2 350

Serial Number: 008871

Test Date: 02/23/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:32pm |
| AIR BLK  | .00    | 9:33pm |
| ACCY CHK | .08    | 9:33pm |
| AIR BLK  | .00    | 9:34pm |
| SUB TEST | .00    | 9:35pm |
| AIR BLK  | .00    | 9:36pm |
| SUB TEST | .00    | 9:37pm |
| AIR BLK  | .00    | 9:38pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**GASTON BAT MOBILE UNIT 2 350**

Serial Number: 008871      Test Record Number: 948  
Test Date: 02/23/2019      Test Time: 9:39pm EST

**System Check: Passed**

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:39pm |
| FLO  | Pass   | 9:39pm |
| FC   | Pass   | 9:39pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:39pm |
| SRC  | Pass   | 9:39pm |
| DET  | Pass   | 9:39pm |
| BAR  | Pass   | 9:39pm |
| BT   | Pass   | 9:39pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:40pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:40pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:40pm |
| CAL  | Pass   | 9:40pm |

**Preventive Maintenance**  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Graham

Instrument Location Graham Co. SO.

Instrument Serial No. 008915

Robbinsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Carter

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GRAHAM COUNTY GRAHAM COUNTY SD 370

Serial Number: 008915

Test Date: 02/19/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734102


Exp Date: 12/07/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:00pm |
| AIR BLK  | .00    | 1:01pm |
| ACCY CHK | .08    | 1:02pm |
| AIR BLK  | .00    | 1:03pm |
| SUB TEST | .00    | 1:04pm |
| AIR BLK  | .00    | 1:05pm |
| SUB TEST | .00    | 1:06pm |
| AIR BLK  | .00    | 1:07pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**GRAHAM COUNTY GRAHAM COUNTY SD 370**

Serial Number: 008915      Test Record Number: 752  
Test Date: 02/19/2019      Test Time: 1:08pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:09pm |
| FLO  | Pass   | 1:09pm |
| FC   | Pass   | 1:09pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:09pm |
| SRC  | Pass   | 1:09pm |
| DET  | Pass   | 1:09pm |
| BAR  | Pass   | 1:09pm |
| BT   | Pass   | 1:09pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:10pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:10pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:10pm |
| CAL  | Pass   | 1:10pm |

Preventive Maintenance  
Status: Pass



**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Greene Instrument Location Greene Co. S.O.  
Instrument Serial No. 008670 301 W. Greene St., Snow Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670  
Test Date: 02/15/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G  
Permit Number: 12955E  
Effective:  
06/01/2017-06/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG807102  
Exp Date: 03/12/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:16am |
| AIR BLK  | .00    | 11:17am |
| ACCY CHK | .08    | 11:18am |
| AIR BLK  | .00    | 11:19am |
| SUB TEST | .00    | 11:19am |
| AIR BLK  | .00    | 11:20am |
| SUB TEST | .00    | 11:22am |
| AIR BLK  | .00    | 11:23am |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670      Test Record Number: 1836  
Test Date: 02/15/2019      Test Time: 11:24am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:24am |
| FLO  | Pass   | 11:24am |
| FC   | Pass   | 11:24am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:24am |
| SRC  | Pass   | 11:24am |
| DET  | Pass   | 11:24am |
| BAR  | Pass   | 11:24am |
| BT   | Pass   | 11:24am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:25am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:25am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:25am |
| CAL  | Pass   | 11:25am |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location High Point  
Instrument Serial No. 008718 Jail

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**GUILFORD COUNTY HIGH POINT JAIL 401**

Serial Number: 008718

Test Date: 02/19/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020


| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 12:01pm        |
| AIR BLK         | .00        | 12:02pm        |
| ACCY CHK        | .07        | 12:03pm        |
| AIR BLK         | .00        | 12:04pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:05pm</b> |
| AIR BLK         | .00        | 12:05pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:07pm</b> |
| AIR BLK         | .00        | 12:08pm        |

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**GUILFORD COUNTY HIGH POINT JAIL 401**

Serial Number: 008718      Test Record Number: 1960  
Test Date: 02/19/2019      Test Time: 12:10pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:10pm |
| FLO  | Pass   | 12:10pm |
| FC   | Pass   | 12:11pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:11pm |
| SRC  | Pass   | 12:11pm |
| DET  | Pass   | 12:11pm |
| BAR  | Pass   | 12:11pm |
| BT   | Pass   | 12:11pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:11pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:11pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:11pm |
| CAL  | Pass   | 12:11pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Haywood

Instrument Location Haywood Co. Jail

Instrument Serial No. 008714

Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul R. Cuth

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008714  
Test Date: 02/25/2019

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734102


Exp Date: 12/07/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:41pm |
| AIR BLK  | .00    | 1:42pm |
| ACCY CHK | .07    | 1:43pm |
| AIR BLK  | .00    | 1:43pm |
| SUB TEST | .00    | 1:44pm |
| AIR BLK  | .00    | 1:45pm |
| SUB TEST | .00    | 1:47pm |
| AIR BLK  | .00    | 1:48pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008714      Test Record Number: 1515  
Test Date: 02/25/2019      Test Time: 1:48pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:48pm |
| FLO  | Pass   | 1:48pm |
| FC   | Pass   | 1:49pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:49pm |
| SRC  | Pass   | 1:49pm |
| DET  | Pass   | 1:49pm |
| BAR  | Pass   | 1:49pm |
| BT   | Pass   | 1:49pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:49pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:49pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:49pm |
| CAL  | Pass   | 1:49pm |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Haywood Instrument Location Haywood Co. Jail  
Instrument Serial No. 008712 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

635  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008712

Test Date: 02/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:40pm |
| AIR BLK  | .00    | 1:41pm |
| ACCY CHK | .08    | 1:41pm |
| AIR BLK  | .00    | 1:42pm |
| SUB TEST | .00    | 1:43pm |
| AIR BLK  | .00    | 1:44pm |
| SUB TEST | .00    | 1:45pm |
| AIR BLK  | .00    | 1:46pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430**

Serial Number: 008712      Test Record Number: 2180  
Test Date: 02/25/2019      Test Time: 1:47pm EST

**System Check: Passed**

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:47pm |
| FLO  | Pass   | 1:47pm |
| FC   | Pass   | 1:48pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:48pm |
| SRC  | Pass   | 1:48pm |
| DET  | Pass   | 1:48pm |
| BAR  | Pass   | 1:48pm |
| BT   | Pass   | 1:48pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:48pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:48pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:48pm |
| CAL  | Pass   | 1:48pm |

**Preventive Maintenance  
Status: Pass**

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson

Instrument Location Henderson Co. Detention

Instrument Serial No. 008822

Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**HENDERSON COUNTY DETENTION 440**

Serial Number: 008822

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 4:09pm        |
| AIR BLK         | .00        | 4:10pm        |
| ACCY CHK        | .08        | 4:11pm        |
| AIR BLK         | .00        | 4:12pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>4:13pm</b> |
| AIR BLK         | .00        | 4:14pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>4:16pm</b> |
| AIR BLK         | .00        | 4:16pm        |

**Reported AC: .00 g/210L**

Signature of Chemical Analyst

Court CVR



**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*HENDERSON COUNTY DETENTION 440*

Serial Number: 008822      Test Record Number: 2317  
Test Date: 02/21/2019      Test Time: 4:18pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 4:19pm |
| FLO  | Pass   | 4:19pm |
| FC   | Pass   | 4:19pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 4:19pm |
| SRC  | Pass   | 4:19pm |
| DET  | Pass   | 4:19pm |
| BAR  | Pass   | 4:19pm |
| BT   | Pass   | 4:19pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 4:19pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 4:20pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 4:20pm |
| CAL  | Pass   | 4:20pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson Instrument Location Henderson Co Detention  
Instrument Serial No. 008806 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**HENDERSON COUNTY DENTENTION 440**

Serial Number: 008806

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: ,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 4:09pm        |
| AIR BLK         | .00        | 4:10pm        |
| ACCY CHK        | .07        | 4:11pm        |
| AIR BLK         | .00        | 4:12pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>4:12pm</b> |
| AIR BLK         | .00        | 4:13pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>4:15pm</b> |
| AIR BLK         | .00        | 4:16pm        |

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*HENDERSON COUNTY DENTENTION 440*

Serial Number: 008806      Test Record Number: 2692

Test Date: 02/21/2019      Test Time: 4:16pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 4:17pm |
| FLO  | Pass   | 4:17pm |
| FC   | Pass   | 4:17pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 4:17pm |
| SRC  | Pass   | 4:17pm |
| DET  | Pass   | 4:17pm |
| BAR  | Pass   | 4:17pm |
| BT   | Pass   | 4:17pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 4:18pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 4:18pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 4:18pm |
| CAL  | Pass   | 4:18pm |

**Preventive Maintenance**

Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Hoke Instrument Location Hoke County  
Instrument Serial No. 008852 DETENTION CENTER  
RAEFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Al R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**HOKE COUNTY DETENTION CENTER 460**

Serial Number: 008852

Test Date: 02/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:30pm |
| AIR BLK  | .00    | 1:30pm |
| ACCY CHK | .08    | 1:31pm |
| AIR BLK  | .00    | 1:32pm |
| SUB TEST | .00    | 1:33pm |
| AIR BLK  | .00    | 1:34pm |
| SUB TEST | .00    | 1:35pm |
| AIR BLK  | .00    | 1:36pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**HOKE COUNTY DETENTION CENTER 460**

Serial Number: 008852      Test Record Number: 903  
Test Date: 02/25/2019      Test Time: 1:37pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:37pm |
| FLO  | Pass   | 1:37pm |
| FC   | Pass   | 1:37pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:37pm |
| SRC  | Pass   | 1:37pm |
| DET  | Pass   | 1:37pm |
| BAR  | Pass   | 1:37pm |
| BT   | Pass   | 1:37pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:38pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:38pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:38pm |
| CAL  | Pass   | 1:38pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Hyde Instrument Location Hyde Co. S.O. - Ocracoke  
Instrument Serial No. 008797 NC 12, Ocracoke, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>nd</sup> day of FEBRUARY, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HYDE COUNTY HYDE CO SO OCRACOE 470

Serial Number: 008797

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:28pm |
| AIR BLK  | .00    | 12:29pm |
| ACCY CHK | .07    | 12:29pm |
| AIR BLK  | .00    | 12:31pm |
| SUB TEST | .00    | 12:31pm |
| AIR BLK  | .00    | 12:32pm |
| SUB TEST | .00    | 12:34pm |
| AIR BLK  | .00    | 12:35pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**HYDE COUNTY HYDE CO SO OCRACOKE 470**

Serial Number: 008797      Test Record Number: 560  
Test Date: 02/22/2019      Test Time: 12:35pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:36pm |
| FLO  | Pass   | 12:36pm |
| FC   | Pass   | 12:36pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:36pm |
| SRC  | Pass   | 12:36pm |
| DET  | Pass   | 12:36pm |
| BAR  | Pass   | 12:36pm |
| BT   | Pass   | 12:36pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:37pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:37pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:37pm |
| CAL  | Pass   | 12:37pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Iredell County SO  
Instrument Serial No. 008809 201 E. Water St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

IREDELL COUNTY IREDELL COUNTY SO 480

Serial Number: 008809  
Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:06am |
| AIR BLK  | .00    | 9:07am |
| ACCY CHK | .08    | 9:08am |
| AIR BLK  | .00    | 9:09am |
| SUB TEST | .00    | 9:09am |
| AIR BLK  | .00    | 9:10am |
| SUB TEST | .00    | 9:12am |
| AIR BLK  | .00    | 9:13am |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

IREDELL COUNTY IREDELL COUNTY SO 480

Serial Number: 008809      Test Record Number: 4095  
Test Date: 02/15/2019      Test Time: 9:14am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:14am |
| FLO  | Pass   | 9:14am |
| FC   | Pass   | 9:14am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:14am |
| SRC  | Pass   | 9:14am |
| DET  | Pass   | 9:14am |
| BAR  | Pass   | 9:14am |
| BT   | Pass   | 9:14am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:15am |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:15am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:15am |
| CAL  | Pass   | 9:15am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Statesville PD  
Instrument Serial No. 008619 300 S. Todd St. Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**IREDELL COUNTY STATESVILLE PD 480**

Serial Number: 008619

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

**PREVENTIVE, MAINTENANCE**

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 9:48am        |
| AIR BLK         | .00        | 9:49am        |
| ACCY CHK        | .07        | 9:50am        |
| AIR BLK         | .00        | 9:51am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:51am</b> |
| AIR BLK         | .00        | 9:52am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:54am</b> |
| AIR BLK         | .00        | 9:55am        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*IREDELL COUNTY STATESVILLE PD 480*

Serial Number: 008619      Test Record Number: 1536  
Test Date: 02/15/2019      Test Time: 9:56am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:56am |
| FLO  | Pass   | 9:56am |
| FC   | Pass   | 9:57am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:57am |
| SRC  | Pass   | 9:57am |
| DET  | Pass   | 9:57am |
| BAR  | Pass   | 9:57am |
| BT   | Pass   | 9:57am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:57am |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:57am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:58am |
| CAL  | Pass   | 9:58am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Iredell County SO  
Instrument Serial No. 008684 201 E. Water St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. D. Ray  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

IREDELL COUNTY IREDELL COUNTY SO 480

Serial Number: 008684

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 8:49am |
| AIR BLK  | .00    | 8:50am |
| ACCY CHK | .08    | 8:50am |
| AIR BLK  | .00    | 8:51am |
| SUB TEST | .00    | 8:52am |
| AIR BLK  | .00    | 8:53am |
| SUB TEST | .00    | 8:54am |
| AIR BLK  | .00    | 8:55am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

IREDELL COUNTY IREDELL COUNTY SO 480

Serial Number: 008684      Test Record Number: 4249  
Test Date: 02/27/2019      Test Time: 8:56am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 8:56am |
| FLO  | Pass   | 8:56am |
| FC   | Pass   | 8:56am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 8:57am |
| SRC  | Pass   | 8:57am |
| DET  | Pass   | 8:57am |
| BAR  | Pass   | 8:57am |
| BT   | Pass   | 8:57am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 8:57am |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 8:57am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 8:57am |
| CAL  | Pass   | 8:57am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County JOHNSTON Instrument Location JOHNSTON COUNTY  
Instrument Serial No. 008810 DETENTION CENTER  
SMITHFIELD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

JOHNSTON COUNTY DETENTION CENTER 500

Serial Number: 008810  
Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202

Exp Date: 01/22/2021

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 10:01am        |
| AIR BLK         | .00        | 10:02am        |
| ACCY CHK        | .08        | 10:03am        |
| AIR BLK         | .00        | 10:04am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>10:04am</b> |
| AIR BLK         | .00        | 10:05am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>10:06am</b> |
| AIR BLK         | .00        | 10:07am        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*JOHNSTON COUNTY DETENTION CENTER 500*

Serial Number: 008810      Test Record Number: 3604  
Test Date: 02/21/2019      Test Time: 10:08am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:08am |
| FLO  | Pass   | 10:08am |
| FC   | Pass   | 10:08am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:08am |
| SRC  | Pass   | 10:08am |
| DET  | Pass   | 10:08am |
| BAR  | Pass   | 10:08am |
| BT   | Pass   | 10:08am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:09am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:09am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:09am |
| CAL  | Pass   | 10:09am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County JOHNSTON

Instrument Location JOHNSTON COUNTY


Instrument Serial No. 008846

DETENTION CENTER  
SMITHFIELD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Rg Bams

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

JOHNSTON COUNTY DETENTION CENTER 500

Serial Number: 008846

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R.

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202


Exp Date: 01/22/2021

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:12am |
| AIR BLK  | .00    | 10:12am |
| ACCY CHK | .08    | 10:13am |
| AIR BLK  | .00    | 10:14am |
| SUB TEST | .00    | 10:15am |
| AIR BLK  | .00    | 10:16am |
| SUB TEST | .00    | 10:18am |
| AIR BLK  | .00    | 10:19am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**JOHNSTON COUNTY DETENTION CENTER 500**

Serial Number: 008846      Test Record Number: 4522  
Test Date: 02/21/2019      Test Time: 10:19am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:19am |
| FLO  | Pass   | 10:19am |
| FC   | Pass   | 10:20am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:20am |
| SRC  | Pass   | 10:20am |
| DET  | Pass   | 10:20am |
| BAR  | Pass   | 10:20am |
| BT   | Pass   | 10:20am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:20am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:20am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:21am |
| CAL  | Pass   | 10:21am |

Preventive Maintenance  
Status: Pass



**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lincoln

Instrument Location Lincoln County Courthouse


Instrument Serial No. 008823


#1 Courthouse Square, Lincoln

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.





Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008823

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG722408


Exp Date: 08/12/2019

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 2:35pm        |
| AIR BLK         | .00        | 2:35pm        |
| ACCY CHK        | .08        | 2:36pm        |
| AIR BLK         | .00        | 2:37pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>2:37pm</b> |
| AIR BLK         | .00        | 2:38pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>2:40pm</b> |
| AIR BLK         | .00        | 2:41pm        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**LINCOLN COUNTY COURTHOUSE 540**

Serial Number: 008823      Test Record Number: 1469  
Test Date: 02/15/2019      Test Time: 2:43pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:43pm |
| FLO  | Pass   | 2:43pm |
| FC   | Pass   | 2:43pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:43pm |
| SRC  | Pass   | 2:43pm |
| DET  | Pass   | 2:43pm |
| BAR  | Pass   | 2:43pm |
| BT   | Pass   | 2:43pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:44pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:44pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:44pm |
| CAL  | Pass   | 2:44pm |

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lincoln Instrument Location Lincoln County Courthouse  
Instrument Serial No. 008827 #1 Court Square, Lincoln

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Holt  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**LINCOLN COUNTY COURTHOUSE 540**

Serial Number: 008827

Test Date: 02/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:20pm |
| AIR BLK  | .00    | 2:21pm |
| ACCY CHK | .07    | 2:22pm |
| AIR BLK  | .00    | 2:23pm |
| SUB TEST | .00    | 2:24pm |
| AIR BLK  | .00    | 2:25pm |
| SUB TEST | .00    | 2:26pm |
| AIR BLK  | .00    | 2:27pm |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008827      Test Record Number: 3075  
Test Date: 02/04/2019      Test Time: 2:28pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:28pm |
| FLO  | Pass   | 2:28pm |
| FC   | Pass   | 2:28pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:28pm |
| SRC  | Pass   | 2:28pm |
| DET  | Pass   | 2:28pm |
| BAR  | Pass   | 2:28pm |
| BT   | Pass   | 2:28pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:29pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:29pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:29pm |
| CAL  | Pass   | 2:29pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Macon Instrument Location Macon Co. Jail  
Instrument Serial No. 008789 Franklin, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuth

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MACON COUNTY MACON COUNTY JAIL 550

Serial Number: 008789

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716202


Exp Date: 06/11/2019

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 9:46am        |
| AIR BLK         | .00        | 9:47am        |
| ACCY CHK        | .07        | 9:47am        |
| AIR BLK         | .00        | 9:48am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:49am</b> |
| AIR BLK         | .00        | 9:50am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:51am</b> |
| AIR BLK         | .00        | 9:52am        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MACON COUNTY MACON COUNTY JAIL 550

Serial Number: 008789      Test Record Number: 637  
Test Date: 02/27/2019      Test Time: 9:53am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:54am |
| FLO  | Pass   | 9:54am |
| FC   | Pass   | 9:54am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:54am |
| SRC  | Pass   | 9:54am |
| DET  | Pass   | 9:54am |
| BAR  | Pass   | 9:54am |
| BT   | Pass   | 9:54am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:54am |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:55am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:55am |
| CAL  | Pass   | 9:55am |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Macon

Instrument Location Macon Co. Jail


Instrument Serial No. 008618

Franklin, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Carol R. Cutler

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MACON COUNTY MACON COUNTY JAIL 550

Serial Number: 008618

Test Date: 02/27/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG722408


Exp Date: 08/12/2019

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 9:42am        |
| AIR BLK         | .00        | 9:43am        |
| ACCY CHK        | .07        | 9:44am        |
| AIR BLK         | .00        | 9:45am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:46am</b> |
| AIR BLK         | .00        | 9:47am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>9:48am</b> |
| AIR BLK         | .00        | 9:49am        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

MACON COUNTY MACON COUNTY JAIL 550

Serial Number: 008618      Test Record Number: 1922  
Test Date: 02/27/2019      Test Time: 9:50am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:51am |
| FLO  | Pass   | 9:51am |
| FC   | Pass   | 9:51am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:51am |
| SRC  | Pass   | 9:51am |
| DET  | Pass   | 9:51am |
| BAR  | Pass   | 9:51am |
| BT   | Pass   | 9:51am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:51am |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:52am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:52am |
| CAL  | Pass   | 9:52am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County MARTIN Instrument Location MARTIN CO. S.O.

Instrument Serial No. 008918 305 E. MAIN ST., WILLIAMSTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26<sup>th</sup> day of FEBRUARY, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008918

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 3:06pm |
| AIR BLK  | .00    | 3:07pm |
| ACCY CHK | .08    | 3:08pm |
| AIR BLK  | .00    | 3:09pm |
| SUB TEST | .00    | 3:09pm |
| AIR BLK  | .00    | 3:10pm |
| SUB TEST | .00    | 3:11pm |
| AIR BLK  | .00    | 3:12pm |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008918      Test Record Number: 665  
Test Date: 02/26/2019      Test Time: 3:03pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 3:03pm |
| FLO  | Pass   | 3:03pm |
| FC   | Pass   | 3:03pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 3:03pm |
| SRC  | Pass   | 3:03pm |
| DET  | Pass   | 3:03pm |
| BAR  | Pass   | 3:03pm |
| BT   | Pass   | 3:03pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 3:04pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 3:04pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 3:04pm |
| CAL  | Pass   | 3:04pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE 3  
Instrument Serial No. 008969 CMRD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBERG BAT MOBILE 3 590

Serial Number: 008969

Test Date: 02/06/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 07281E

Effective:

02/01/2018-02/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 9:58pm  |
| AIR BLK  | .00    | 9:59pm  |
| ACCY CHK | .08    | 9:59pm  |
| AIR BLK  | .00    | 10:00pm |
| SUB TEST | .00    | 10:01pm |
| AIR BLK  | .00    | 10:01pm |
| SUB TEST | .00    | 10:04pm |
| AIR BLK  | .00    | 10:05pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**MECKLENBERG BAT MOBILE 3 590**

Serial Number: 008969      Test Record Number: 263  
Test Date: 02/06/2019      Test Time: 9:47pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:47pm |
| FLO  | Pass   | 9:47pm |
| FC   | Pass   | 9:47pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:48pm |
| SRC  | Pass   | 9:48pm |
| DET  | Pass   | 9:48pm |
| BAR  | Pass   | 9:48pm |
| BT   | Pass   | 9:48pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:48pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:48pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:48pm |
| CAL  | Pass   | 9:48pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Cornelius PD  
Instrument Serial No. 008692 21440 Catawba Ave., Cornelius

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:47pm |
| AIR BLK  | .00    | 12:48pm |
| ACCY CHK | .08    | 12:49pm |
| AIR BLK  | .00    | 12:50pm |
| SUB TEST | .00    | 12:51pm |
| AIR BLK  | .00    | 12:52pm |
| SUB TEST | .00    | 12:53pm |
| AIR BLK  | .00    | 12:54pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692

Test Record Number: 2872

Test Date: 02/15/2019

Test Time: 12:56pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:57pm |
| FLO  | Pass   | 12:57pm |
| FC   | Pass   | 12:57pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:57pm |
| SRC  | Pass   | 12:57pm |
| DET  | Pass   | 12:57pm |
| BAR  | Pass   | 12:57pm |
| BT   | Pass   | 12:57pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:58pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:58pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:58pm |
| CAL  | Pass   | 12:58pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location CMRD-LEC  
Instrument Serial No. 008691 600 E. Trade St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. D. [Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**MECKLENBURG COUNTY CMPD LEC 590**

Serial Number: 008691

Test Date: 02/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 11:37am        |
| AIR BLK         | .00        | 11:38am        |
| ACCY CHK        | .07        | 11:39am        |
| AIR BLK         | .00        | 11:40am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:41am</b> |
| AIR BLK         | .00        | 11:42am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:43am</b> |
| AIR BLK         | .00        | 11:44am        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008691      Test Record Number: 7473  
Test Date: 02/12/2019      Test Time: 11:45am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:45am |
| FLO  | Pass   | 11:45am |
| FC   | Pass   | 11:45am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:46am |
| SRC  | Pass   | 11:46am |
| DET  | Pass   | 11:46am |
| BAR  | Pass   | 11:46am |
| BT   | Pass   | 11:46am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:46am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:46am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:46am |
| CAL  | Pass   | 11:46am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Mecklenburg Instrument Location Bad Mobile Unit 2

Instrument Serial No. 008671

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

658  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG BAT MOBILE UNIT 2 590

Serial Number: 008871

Test Date: 02/13/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 6:42pm        |
| AIR BLK         | .00        | 6:43pm        |
| ACCY CHK        | .08        | 6:43pm        |
| AIR BLK         | .00        | 6:44pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>6:45pm</b> |
| AIR BLK         | .00        | 6:46pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>6:47pm</b> |
| AIR BLK         | .00        | 6:48pm        |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**MECKLENBURG BAT MOBILE UNIT 2 590**

Serial Number: 008871      Test Record Number: 945

Test Date: 02/13/2019      Test Time: 6:49pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 6:49pm |
| FLO  | Pass   | 6:49pm |
| FC   | Pass   | 6:49pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 6:49pm |
| SRC  | Pass   | 6:49pm |
| DET  | Pass   | 6:49pm |
| BAR  | Pass   | 6:49pm |
| BT   | Pass   | 6:49pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 6:50pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 6:50pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 6:50pm |
| CAL  | Pass   | 6:50pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County MONTGOMERY Instrument Location MONTGOMERY COUNTY  
Instrument Serial No. 008709 DETENTION CENTER  
TRAY, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bano

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**MONTGOMERY COUNTY DETENTION CENTER 610**

Serial Number: 008709

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 12:39pm        |
| AIR BLK         | .00        | 12:39pm        |
| ACCY CHK        | .07        | 12:40pm        |
| AIR BLK         | .00        | 12:41pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:42pm</b> |
| AIR BLK         | .00        | 12:43pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>12:44pm</b> |
| AIR BLK         | .00        | 12:45pm        |

**Reported AC: .00 g/210L**

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**MONTGOMERY COUNTY DETENTION CENTER 610**

Serial Number: 008709      Test Record Number: 1150

Test Date: 02/26/2019      Test Time: 12:46pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:46pm |
| FLO  | Pass   | 12:46pm |
| FC   | Pass   | 12:46pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:46pm |
| SRC  | Pass   | 12:46pm |
| DET  | Pass   | 12:46pm |
| BAR  | Pass   | 12:46pm |
| BT   | Pass   | 12:46pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:47pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:47pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:47pm |
| CAL  | Pass   | 12:47pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County MONTGOMERY Instrument Location MONTGOMERY COUNTY

Instrument Serial No. 008657 DETENTION CENTER  
TRCY, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Al R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MONTGOMERY COUNTY DETENTION CENTER 610

Serial Number: 008657

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:36pm |
| AIR BLK  | .00    | 12:36pm |
| ACCY CHK | .08    | 12:37pm |
| AIR BLK  | .00    | 12:38pm |
| SUB TEST | .00    | 12:39pm |
| AIR BLK  | .00    | 12:40pm |
| SUB TEST | .00    | 12:41pm |
| AIR BLK  | .00    | 12:42pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**MONTGOMERY COUNTY DETENTION CENTER 610**

Serial Number: 008657      Test Record Number: 1593  
Test Date: 02/26/2019      Test Time: 12:43pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:43pm |
| FLO  | Pass   | 12:43pm |
| FC   | Pass   | 12:43pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:43pm |
| SRC  | Pass   | 12:43pm |
| DET  | Pass   | 12:43pm |
| BAR  | Pass   | 12:43pm |
| BT   | Pass   | 12:43pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:44pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:44pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:44pm |
| CAL  | Pass   | 12:44pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County NASH Instrument Location Nashville PD

Instrument Serial No. 008630 501 S. BARNES ST NASHVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Adrian Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NASH COUNTY NASHVILLE PD 630

Serial Number: 008630  
Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902106


Exp Date: 01/21/2021

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 3:06pm |
| AIR BLK  | .00    | 3:06pm |
| ACCY CHK | .08    | 3:07pm |
| AIR BLK  | .00    | 3:08pm |
| SUB TEST | .00    | 3:09pm |
| AIR BLK  | .00    | 3:10pm |
| SUB TEST | .00    | 3:12pm |
| AIR BLK  | .00    | 3:13pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

NASH COUNTY NASHVILLE PD 630

Serial Number: 008630      Test Record Number: 4363  
Test Date: 02/28/2019      Test Time: 3:13pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 3:13pm |
| FLO  | Pass   | 3:13pm |
| FC   | Pass   | 3:14pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 3:14pm |
| SRC  | Pass   | 3:14pm |
| DET  | Pass   | 3:14pm |
| BAR  | Pass   | 3:14pm |
| BT   | Pass   | 3:14pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 3:14pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 3:14pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 3:15pm |
| CAL  | Pass   | 3:15pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ORANGE

Instrument Location Hillsborough PD


Instrument Serial No. 008924

127 N. Clinton St. Hillsborough, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008924

Test Date: 02/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: NO

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 5:15pm |
| AIR BLK  | .00    | 5:16pm |
| ACCY CHK | .07    | 5:16pm |
| AIR BLK  | .00    | 5:17pm |
| SUB TEST | .00    | 5:18pm |
| AIR BLK  | .00    | 5:19pm |
| SUB TEST | .00    | 5:21pm |
| AIR BLK  | .00    | 5:22pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008924      Test Record Number: 1408  
Test Date: 02/12/2019      Test Time: 5:24pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 5:24pm |
| FLO  | Pass   | 5:24pm |
| FC   | Pass   | 5:24pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 5:24pm |
| SRC  | Pass   | 5:24pm |
| DET  | Pass   | 5:24pm |
| BAR  | Pass   | 5:24pm |
| BT   | Pass   | 5:24pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 5:25pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 5:25pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 5:25pm |
| CAL  | Pass   | 5:25pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ORANGE

Instrument Location Hillsborough PD


Instrument Serial No. 008873

127 N. Chatham St Hillsborough, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008873

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:25am |
| AIR BLK  | .00    | 9:26am |
| ACCY CHK | .07    | 9:26am |
| AIR BLK  | .00    | 9:27am |
| SUB TEST | .00    | 9:28am |
| AIR BLK  | .00    | 9:29am |
| SUB TEST | .00    | 9:30am |
| AIR BLK  | .00    | 9:31am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008873      Test Record Number: 1525  
Test Date: 02/21/2019      Test Time: 9:32am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 9:33am |
| FLO  | Pass   | 9:33am |
| FC   | Pass   | 9:33am |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 9:33am |
| SRC  | Pass   | 9:33am |
| DET  | Pass   | 9:33am |
| BAR  | Pass   | 9:33am |
| BT   | Pass   | 9:33am |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 9:34am |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 9:34am |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 9:34am |
| CAL  | Pass   | 9:34am |

Preventive Maintenance  
Status: *Pass*

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pender Instrument Location Pender County  
Instrument Serial No. 008948 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**PENDER COUNTY SHERIFF DEPT ANNEX 700**

Serial Number: 008948

Test Date: 02/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:18am |
| AIR BLK  | .00    | 11:19am |
| ACCY CHK | .07    | 11:20am |
| AIR BLK  | .00    | 11:20am |
| SUB TEST | .00    | 11:21am |
| AIR BLK  | .00    | 11:22am |
| SUB TEST | .00    | 11:23am |
| AIR BLK  | .00    | 11:24am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**PENDER COUNTY SHERIFF DEPT ANNEX 700**

Serial Number: 008948      Test Record Number: 843  
Test Date: 02/01/2019      Test Time: 11:25am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:25am |
| FLO  | Pass   | 11:25am |
| FC   | Pass   | 11:25am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:25am |
| SRC  | Pass   | 11:25am |
| DET  | Pass   | 11:25am |
| BAR  | Pass   | 11:25am |
| BT   | Pass   | 11:25am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:26am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:26am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:26am |
| CAL  | Pass   | 11:26am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Ayden P.D.  
Instrument Serial No. 008666 4144 West Ave., Ayden, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT AYDEN PD 730

Serial Number: 008666

Test Date: 02/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:17pm |
| AIR BLK  | .00    | 1:18pm |
| ACCY CHK | .07    | 1:18pm |
| AIR BLK  | .00    | 1:19pm |
| SUB TEST | .00    | 1:20pm |
| AIR BLK  | .00    | 1:21pm |
| SUB TEST | .00    | 1:22pm |
| AIR BLK  | .00    | 1:23pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

PITT AYDEN PD 730

Serial Number: 008666      Test Record Number: 1072  
Test Date: 02/12/2019      Test Time: 1:24pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:25pm |
| FLO  | Pass   | 1:25pm |
| FC   | Pass   | 1:25pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:25pm |
| SRC  | Pass   | 1:25pm |
| DET  | Pass   | 1:25pm |
| BAR  | Pass   | 1:25pm |
| BT   | Pass   | 1:25pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:26pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:26pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:26pm |
| CAL  | Pass   | 1:26pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Polk Instrument Location Polk County LEC  
Instrument Serial No. 0088881 880 ENC 108, Columbus

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

POLK COUNTY POLK COUNTY LEC 740

Serial Number: 008881

Test Date: 02/20/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:17am |
| AIR BLK  | .00    | 10:18am |
| ACCY CHK | .08    | 10:18am |
| AIR BLK  | .00    | 10:20am |
| SUB TEST | .00    | 10:20am |
| AIR BLK  | .00    | 10:21am |
| SUB TEST | .00    | 10:23am |
| AIR BLK  | .00    | 10:24am |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

POLK COUNTY POLK COUNTY LEC 740

Serial Number: 008881      Test Record Number: 795  
Test Date: 02/20/2019      Test Time: 10:25am EDT

System Check: Passed

Baseline Tests

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:26am |
| FLO  | Pass   | 10:26am |
| FC   | Pass   | 10:26am |

Temperature Tests

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:26am |
| SRC  | Pass   | 10:26am |
| DET  | Pass   | 10:26am |
| BAR  | Pass   | 10:26am |
| BT   | Pass   | 10:26am |

Blank Tests

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:27am |


Printer Tests

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:27am |

CRC Tests

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:27am |
| CAL  | Pass   | 10:27am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Randolph Co. Instrument Location Randolph Police Dept.  
Instrument Serial No. 008137 Randolph, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28<sup>th</sup> day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



A handwritten signature in black ink, appearing to be "D. P. Kunk", written over a horizontal line.

Signature of Certifying Official

654

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY RANDLEMAN PD 750

Serial Number: 008737

Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805802


Exp Date: 02/27/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:08am |
| AIR BLK  | .00    | 11:09am |
| ACCY CHK | .08    | 11:10am |
| AIR BLK  | .00    | 11:11am |
| SUB TEST | .00    | 11:11am |
| AIR BLK  | .00    | 11:12am |
| SUB TEST | .00    | 11:14am |
| AIR BLK  | .00    | 11:15am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**RANDOLPH COUNTY RANDLEMAN PD 750**

Serial Number: 008737      Test Record Number: 1077  
Test Date: 02/28/2019      Test Time: 11:15am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:15am |
| FLO  | Pass   | 11:15am |
| FC   | Pass   | 11:16am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:16am |
| SRC  | Pass   | 11:16am |
| DET  | Pass   | 11:16am |
| BAR  | Pass   | 11:16am |
| BT   | Pass   | 11:16am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:16am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:16am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:16am |
| CAL  | Pass   | 11:16am |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Randolph Co. Instrument Location Randolph Co. Detention Center

Instrument Serial No. 008860 Asheboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of FEBRUARY, 20 28, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008860

Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 9:55am  |
| AIR BLK  | .00    | 9:56am  |
| ACCY CHK | .08    | 9:56am  |
| AIR BLK  | .00    | 9:57am  |
| SUB TEST | .00    | 9:58am  |
| AIR BLK  | .00    | 9:59am  |
| SUB TEST | .00    | 10:01am |
| AIR BLK  | .00    | 10:01am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*RANDOLPH COUNTY DETENTION CENTER 750*

Serial Number: 008860      Test Record Number: 2672  
Test Date: 02/28/2019      Test Time: 10:06am EST

System Check: *Passed*

Baseline Tests

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:06am |
| FLO  | Pass   | 10:06am |
| FC   | Pass   | 10:06am |

Temperature Tests

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:07am |
| SRC  | Pass   | 10:07am |
| DET  | Pass   | 10:07am |
| BAR  | Pass   | 10:07am |
| BT   | Pass   | 10:07am |

Blank Tests

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:07am |

Printer Tests


| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:07am |

CRC Tests

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:07am |
| CAL  | Pass   | 10:07am |

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Randolph Co. Instrument Location Randolph Co. Detention Center  
Instrument Serial No. 008899 Asheboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**RANDOLPH COUNTY DETENTION CENTER 750**

Serial Number: 008899

Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 9:52am |
| AIR BLK  | .00    | 9:53am |
| ACCY CHK | .08    | 9:54am |
| AIR BLK  | .00    | 9:55am |
| SUB TEST | .00    | 9:55am |
| AIR BLK  | .00    | 9:56am |
| SUB TEST | .00    | 9:58am |
| AIR BLK  | .00    | 9:59am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*RANDOLPH COUNTY DETENTION CENTER 750*

Serial Number: 008899      Test Record Number: 3049  
Test Date: 02/28/2019      Test Time: 10:00am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:00am |
| FLO  | Pass   | 10:00am |
| FC   | Pass   | 10:00am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:00am |
| SRC  | Pass   | 10:00am |
| DET  | Pass   | 10:00am |
| BAR  | Pass   | 10:00am |
| BT   | Pass   | 10:00am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:01am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:01am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:01am |
| CAL  | Pass   | 10:01am |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON

Instrument Location PEMBROKE POLICE DEPT.


Instrument Serial No. 008837

PEMBROKE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Bono

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY PEMBROKE POLICE DEPT  
770

Serial Number: 008837

Test Date: 02/20/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 4:09pm |
| AIR BLK  | .00    | 4:10pm |
| ACCY CHK | .08    | 4:10pm |
| AIR BLK  | .00    | 4:11pm |
| SUB TEST | .00    | 4:12pm |
| AIR BLK  | .00    | 4:13pm |
| SUB TEST | .00    | 4:14pm |
| AIR BLK  | .00    | 4:15pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ROBESON COUNTY PEMBROKE POLICE DEPT 770

Serial Number: 008837      Test Record Number: 1018  
Test Date: 02/20/2019      Test Time: 4:16pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 4:16pm |
| FLO  | Pass   | 4:16pm |
| FC   | Pass   | 4:16pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 4:16pm |
| SRC  | Pass   | 4:16pm |
| DET  | Pass   | 4:16pm |
| BAR  | Pass   | 4:16pm |
| BT   | Pass   | 4:16pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 4:17pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 4:17pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 4:17pm |
| CAL  | Pass   | 4:17pm |

Preventive Maintenance  
Status: Pass

Al Rg B  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location ST PAULS PD  
Instrument Serial No. 008814 ST PAULS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bowers  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY ST PAULS PD 770

Serial Number: 008814

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805802


Exp Date: 02/27/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:18am |
| AIR BLK  | .00    | 10:19am |
| ACCY CHK | .08    | 10:19am |
| AIR BLK  | .00    | 10:21am |
| SUB TEST | .00    | 10:21am |
| AIR BLK  | .00    | 10:22am |
| SUB TEST | .00    | 10:24am |
| AIR BLK  | .00    | 10:25am |

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**ROBESON COUNTY ST PAULS PD 770**

Serial Number: 008814      Test Record Number: 654  
Test Date: 02/22/2019      Test Time: 10:26am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:26am |
| FLO  | Pass   | 10:26am |
| FC   | Pass   | 10:26am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:26am |
| SRC  | Pass   | 10:26am |
| DET  | Pass   | 10:26am |
| BAR  | Pass   | 10:26am |
| BT   | Pass   | 10:26am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:27am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:27am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:27am |
| CAL  | Pass   | 10:27am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ROBESON Instrument Location LUMBERTON PD  
Instrument Serial No. 008629 LUMBERTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY LUMBERTON PD 770

Serial Number: 008629

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:56am |
| AIR BLK  | .00    | 11:57am |
| ACCY CHK | .07    | 11:58am |
| AIR BLK  | .00    | 11:59am |
| SUB TEST | .00    | 12:00pm |
| AIR BLK  | .00    | 12:00pm |
| SUB TEST | .00    | 12:02pm |
| AIR BLK  | .00    | 12:03pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**ROBESON COUNTY LUMBERTON PD 770**

Serial Number: 008629      Test Record Number: 708  
Test Date: 02/22/2019      Test Time: 12:04pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:04pm |
| FLO  | Pass   | 12:04pm |
| FC   | Pass   | 12:04pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:04pm |
| SRC  | Pass   | 12:04pm |
| DET  | Pass   | 12:04pm |
| BAR  | Pass   | 12:04pm |
| BT   | Pass   | 12:04pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:05pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:05pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:05pm |
| CAL  | Pass   | 12:05pm |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location ROBESON COUNTY  
Instrument Serial No. 008836 DETENTION CENTER  
LUMBERTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY DETENTION CENTER 770

Serial Number: 008836

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:52pm |
| AIR BLK  | .00    | 1:53pm |
| ACCY CHK | .08    | 1:53pm |
| AIR BLK  | .00    | 1:54pm |
| SUB TEST | .00    | 1:55pm |
| AIR BLK  | .00    | 1:56pm |
| SUB TEST | .00    | 1:57pm |
| AIR BLK  | .00    | 1:58pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**ROBESON COUNTY DETENTION CENTER 770**

Serial Number: 008836      Test Record Number: 5061  
Test Date: 02/22/2019      Test Time: 1:59pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:59pm |
| FLO  | Pass   | 1:59pm |
| FC   | Pass   | 1:59pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:59pm |
| SRC  | Pass   | 1:59pm |
| DET  | Pass   | 1:59pm |
| BAR  | Pass   | 1:59pm |
| BT   | Pass   | 1:59pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:00pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:00pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:00pm |
| CAL  | Pass   | 2:00pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ROBESON Instrument Location ROBESON COUNTY  
Instrument Serial No. 008805 DETENTION CENTER  
LUMBERTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of FEBRUARY 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY DETENTION CENTER 770

Serial Number: 008805  
Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:54pm |
| AIR BLK  | .00    | 1:55pm |
| ACCY CHK | .08    | 1:55pm |
| AIR BLK  | .00    | 1:56pm |
| SUB TEST | .00    | 1:57pm |
| AIR BLK  | .00    | 1:58pm |
| SUB TEST | .00    | 1:59pm |
| AIR BLK  | .00    | 2:00pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**ROBESON COUNTY DETENTION CENTER 770**

Serial Number: 008805      Test Record Number: 4323  
Test Date: 02/22/2019      Test Time: 2:03pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:04pm |
| FLO  | Pass   | 2:04pm |
| FC   | Pass   | 2:04pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:04pm |
| SRC  | Pass   | 2:04pm |
| DET  | Pass   | 2:04pm |
| BAR  | Pass   | 2:04pm |
| BT   | Pass   | 2:04pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:05pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:05pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:05pm |
| CAL  | Pass   | 2:05pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County ROBESON Instrument Location RED SPRINGS PD  
Instrument Serial No. 008857 RED SPRINGS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY RED SPRINGS PD 770

Serial Number: 008857

Test Date: 02/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:32am |
| AIR BLK  | .00    | 11:33am |
| ACCY CHK | .08    | 11:34am |
| AIR BLK  | .00    | 11:36am |
| SUB TEST | .00    | 11:37am |
| AIR BLK  | .00    | 11:37am |
| SUB TEST | .00    | 11:39am |
| AIR BLK  | .00    | 11:40am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ROBESON COUNTY RED SPRINGS PD 770

Serial Number: 008857      Test Record Number: 547  
Test Date: 02/25/2019      Test Time: 11:41am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:41am |
| FLO  | Pass   | 11:41am |
| FC   | Pass   | 11:41am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:41am |
| SRC  | Pass   | 11:41am |
| DET  | Pass   | 11:41am |
| BAR  | Pass   | 11:41am |
| BT   | Pass   | 11:41am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:42am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:42am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:42am |
| CAL  | Pass   | 11:42am |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Reidsville  
Instrument Serial No. 008784 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Deen  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROCKINGHAM COUNTY REIDSVILLE PD 780

Serial Number: 008784

Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:30pm |
| AIR BLK  | .00    | 2:31pm |
| ACCY CHK | .08    | 2:32pm |
| AIR BLK  | .00    | 2:33pm |
| SUB TEST | .00    | 2:33pm |
| AIR BLK  | .00    | 2:35pm |
| SUB TEST | .00    | 2:36pm |
| AIR BLK  | .00    | 2:37pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ROCKINGHAM COUNTY REIDSVILLE PD 780

Serial Number: 008784      Test Record Number: 1103  
Test Date: 02/18/2019      Test Time: 2:38pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:38pm |
| FLO  | Pass   | 2:38pm |
| FC   | Pass   | 2:38pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:38pm |
| SRC  | Pass   | 2:38pm |
| DET  | Pass   | 2:38pm |
| BAR  | Pass   | 2:38pm |
| BT   | Pass   | 2:38pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:39pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:39pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:39pm |
| CAL  | Pass   | 2:39pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Eden  
Instrument Serial No. 008636 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: 008636

Test Date: 02/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:25pm |
| AIR BLK  | .00    | 12:25pm |
| ACCY CHK | .08    | 12:26pm |
| AIR BLK  | .00    | 12:28pm |
| SUB TEST | .00    | 12:28pm |
| AIR BLK  | .00    | 12:29pm |
| SUB TEST | .00    | 12:30pm |
| AIR BLK  | .00    | 12:31pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: 008636      Test Record Number: 1909  
Test Date: 02/18/2019      Test Time: 12:32pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:32pm |
| FLO  | Pass   | 12:32pm |
| FC   | Pass   | 12:33pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:33pm |
| SRC  | Pass   | 12:33pm |
| DET  | Pass   | 12:33pm |
| BAR  | Pass   | 12:33pm |
| BT   | Pass   | 12:33pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:33pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:33pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:33pm |
| CAL  | Pass   | 12:34pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Rockingham Co Jail  
Instrument Serial No. 008796 Wentworth, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL  
780

Serial Number: 008796  
Test Date: 02/18/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: DEAN, L K  
Permit Number: 11598E  
Effective:  
04/01/2017-04/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG807101  
Exp Date: 03/12/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:37am |
| AIR BLK  | .00    | 10:37am |
| ACCY CHK | .08    | 10:38am |
| AIR BLK  | .00    | 10:39am |
| SUB TEST | .00    | 10:40am |
| AIR BLK  | .00    | 10:41am |
| SUB TEST | .00    | 10:43am |
| AIR BLK  | .00    | 10:44am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL 780

Serial Number: 008796      Test Record Number: 2744  
Test Date: 02/18/2019      Test Time: 10:46am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:46am |
| FLO  | Pass   | 10:46am |
| FC   | Pass   | 10:46am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:46am |
| SRC  | Pass   | 10:46am |
| DET  | Pass   | 10:46am |
| BAR  | Pass   | 10:46am |
| BT   | Pass   | 10:46am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:47am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:47am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:47am |
| CAL  | Pass   | 10:47am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rockingham Instrument Location Madison  
Instrument Serial No. 008802 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: 008802

Test Date: 02/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:01pm |
| AIR BLK  | .00    | 12:02pm |
| ACCY CHK | .08    | 12:02pm |
| AIR BLK  | .00    | 12:04pm |
| SUB TEST | .00    | 12:05pm |
| AIR BLK  | .00    | 12:05pm |
| SUB TEST | .00    | 12:07pm |
| AIR BLK  | .00    | 12:08pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: 008802      Test Record Number: 813  
Test Date: 02/25/2019      Test Time: 12:09pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:10pm |
| FLO  | Pass   | 12:10pm |
| FC   | Pass   | 12:10pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:10pm |
| SRC  | Pass   | 12:10pm |
| DET  | Pass   | 12:10pm |
| BAR  | Pass   | 12:10pm |
| BT   | Pass   | 12:10pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:11pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:11pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:11pm |
| CAL  | Pass   | 12:11pm |

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rowan Instrument Location China Grove  
Instrument Serial No. 008610 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



C. Lee Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROWAN COUNTY CHINA GROVE PD 790

Serial Number: 008610  
Test Date: 02/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:16pm |
| AIR BLK  | .00    | 2:17pm |
| ACCY CHK | .08    | 2:18pm |
| AIR BLK  | .00    | 2:19pm |
| SUB TEST | .00    | 2:20pm |
| AIR BLK  | .00    | 2:21pm |
| SUB TEST | .00    | 2:22pm |
| AIR BLK  | .00    | 2:23pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROWAN COUNTY CHINA GROVE PD 790

Serial Number: 008610      Test Record Number: 2015  
Test Date: 02/04/2019      Test Time: 2:24pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:24pm |
| FLO  | Pass   | 2:24pm |
| FC   | Pass   | 2:24pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:24pm |
| SRC  | Pass   | 2:24pm |
| DET  | Pass   | 2:24pm |
| BAR  | Pass   | 2:24pm |
| BT   | Pass   | 2:24pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:25pm |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:25pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:25pm |
| CAL  | Pass   | 2:25pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rutherford Instrument Location Rutherford County, SO  
Instrument Serial No. 008914 400 N. Washington St. Rutherfordton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RUTHERFORD COUNTY RUTHERFORD COUNTY SO  
800

Serial Number: 008914  
Test Date: 02/08/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:24pm |
| AIR BLK  | .00    | 2:24pm |
| ACCY CHK | .07    | 2:25pm |
| AIR BLK  | .00    | 2:26pm |
| SUB TEST | .00    | 2:27pm |
| AIR BLK  | .00    | 2:27pm |
| SUB TEST | .00    | 2:29pm |
| AIR BLK  | .00    | 2:30pm |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**RUTHERFORD COUNTY RUTHERFORD COUNTY SO 800**

Serial Number: 008914      Test Record Number: 2054  
Test Date: 02/08/2019      Test Time: 2:31pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:31pm |
| FLO  | Pass   | 2:31pm |
| FC   | Pass   | 2:31pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:31pm |
| SRC  | Pass   | 2:31pm |
| DET  | Pass   | 2:31pm |
| BAR  | Pass   | 2:31pm |
| BT   | Pass   | 2:31pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:32pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:32pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:32pm |
| CAL  | Pass   | 2:32pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County SAMPSON Instrument Location SAMPSON COUNTY  
Instrument Serial No. 008825 SHERIFF'S OFFICE  
CLINTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of FEBRUARY, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Benson  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: 008825

Test Date: 02/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202


Exp Date: 01/22/2021

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 5:55pm        |
| AIR BLK         | .00        | 5:56pm        |
| ACCY CHK        | .08        | 5:57pm        |
| AIR BLK         | .00        | 5:58pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>5:59pm</b> |
| AIR BLK         | .00        | 6:00pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>6:01pm</b> |
| AIR BLK         | .00        | 6:02pm        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*SAMPSON COUNTY SAMPSON COUNTY SD 810*

Serial Number: 008825      Test Record Number: 2620  
Test Date: 02/15/2019      Test Time: 6:03pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 6:03pm |
| FLO  | Pass   | 6:03pm |
| FC   | Pass   | 6:04pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 6:04pm |
| SRC  | Pass   | 6:04pm |
| DET  | Pass   | 6:04pm |
| BAR  | Pass   | 6:04pm |
| BT   | Pass   | 6:04pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 6:04pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 6:04pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 6:05pm |
| CAL  | Pass   | 6:05pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County SCOTLAND Instrument Location LAURINBURG PD

Instrument Serial No. 008834 LAURINBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Rg Bano  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SCOTLAND COUNTY LAURINBURG PD 820

Serial Number: 008834

Test Date: 02/20/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG831801

Exp Date: 11/14/2020

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 11:45am        |
| AIR BLK         | .00        | 11:45am        |
| ACCY CHK        | .08        | 11:46am        |
| AIR BLK         | .00        | 11:47am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:48am</b> |
| AIR BLK         | .00        | 11:49am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:50am</b> |
| AIR BLK         | .00        | 11:51am        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

SCOTLAND COUNTY LAURINBURG PD 820

Serial Number: 008834      Test Record Number: 889  
Test Date: 02/20/2019      Test Time: 11:53am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:53am |
| FLO  | Pass   | 11:53am |
| FC   | Pass   | 11:53am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:53am |
| SRC  | Pass   | 11:53am |
| DET  | Pass   | 11:53am |
| BAR  | Pass   | 11:53am |
| BT   | Pass   | 11:53am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:54am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:54am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:54am |
| CAL  | Pass   | 11:54am |

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County SCOTLAND Instrument Location SCOTLAND Co

Instrument Serial No. 008861 SHERIFF'S OFFICE  
LAURINBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of FEBRUARY, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861

Test Date: 02/20/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG831801


Exp Date: 11/14/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:45pm |
| AIR BLK  | .00    | 1:46pm |
| ACCY CHK | .07    | 1:46pm |
| AIR BLK  | .00    | 1:47pm |
| SUB TEST | .00    | 1:48pm |
| AIR BLK  | .00    | 1:49pm |
| SUB TEST | .00    | 1:50pm |
| AIR BLK  | .00    | 1:51pm |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861      Test Record Number: 1493  
Test Date: 02/20/2019      Test Time: 1:52pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:52pm |
| FLO  | Pass   | 1:52pm |
| FC   | Pass   | 1:52pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:52pm |
| SRC  | Pass   | 1:52pm |
| DET  | Pass   | 1:52pm |
| BAR  | Pass   | 1:52pm |
| BT   | Pass   | 1:52pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:53pm |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:53pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:53pm |
| CAL  | Pass   | 1:53pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Locust PD  
Instrument Serial No. 008706 186 Ray Kennedy Drive, Locust

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hatt  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY LOCUST PD 830

Serial Number: 008706

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801


Exp Date: 02/27/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 4:19pm |
| AIR BLK  | .00    | 4:20pm |
| ACCY CHK | .08    | 4:20pm |
| AIR BLK  | .00    | 4:21pm |
| SUB TEST | .00    | 4:22pm |
| AIR BLK  | .00    | 4:23pm |
| SUB TEST | .00    | 4:25pm |
| AIR BLK  | .00    | 4:26pm |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*STANLY LOCUST PD 830*

Serial Number: 008706      Test Record Number: 3516  
Test Date: 02/22/2019      Test Time: 4:27pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 4:27pm |
| FLO  | Pass   | 4:27pm |
| FC   | Pass   | 4:27pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 4:27pm |
| SRC  | Pass   | 4:27pm |
| DET  | Pass   | 4:27pm |
| BAR  | Pass   | 4:27pm |
| BT   | Pass   | 4:27pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 4:28pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 4:28pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 4:28pm |
| CAL  | Pass   | 4:28pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Stanly Co SO  
Instrument Serial No. 008824 126 S. 3rd Street, Albemarle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of February, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hatt  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008824

Test Date: 02/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:53pm |
| AIR BLK  | .00    | 2:54pm |
| ACCY CHK | .08    | 2:55pm |
| AIR BLK  | .00    | 2:56pm |
| SUB TEST | .00    | 2:57pm |
| AIR BLK  | .00    | 2:57pm |
| SUB TEST | .00    | 2:59pm |
| AIR BLK  | .00    | 3:00pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*STANLY COUNTY STANLY COUNTY SD 830*

Serial Number: 008824      Test Record Number: 1396  
Test Date: 02/22/2019      Test Time: 3:01pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 3:01pm |
| FLO  | Pass   | 3:01pm |
| FC   | Pass   | 3:01pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 3:01pm |
| SRC  | Pass   | 3:01pm |
| DET  | Pass   | 3:01pm |
| BAR  | Pass   | 3:01pm |
| BT   | Pass   | 3:01pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 3:02pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 3:02pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 3:02pm |
| CAL  | Pass   | 3:02pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location Stokes County Jail  
Instrument Serial No. 008596 Danbury, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Preventive Maintenance**

**STOKES COUNTY STOKES COUNTY JAIL 840**

Serial Number: 008596      Test Record Number: 1066  
Test Date: 02/25/2019      Test Time: 3:09pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 3:09pm |
| FLO  | Pass   | 3:09pm |
| FC   | Pass   | 3:10pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 3:10pm |
| SRC  | Pass   | 3:10pm |
| DET  | Pass   | 3:10pm |
| BAR  | Pass   | 3:10pm |
| BT   | Pass   | 3:10pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 3:10pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 3:10pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 3:10pm |
| CAL  | Pass   | 3:10pm |

Preventive Maintenance  
Status: Pass



**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**STOKES COUNTY STOKES COUNTY JAIL 840**

Serial Number: 008596      Test Record Number: 1066  
Test Date: 02/25/2019      Test Time: 3:09pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 3:09pm |
| FLO  | Pass   | 3:09pm |
| FC   | Pass   | 3:10pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 3:10pm |
| SRC  | Pass   | 3:10pm |
| DET  | Pass   | 3:10pm |
| BAR  | Pass   | 3:10pm |
| BT   | Pass   | 3:10pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 3:10pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 3:10pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 3:10pm |
| CAL  | Pass   | 3:10pm |

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Surry Instrument Location Elkin Police  
Instrument Serial No. 008926 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Julie Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

*SURRY COUNTY ELKIN PD 850*

Serial Number: 008926

Test Date: 02/12/2019

Citation Number: M0000000-0

Subject's Name:

*PREVENTIVE, MAINTENANCE*

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734102



Exp Date: 12/07/2019

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 1:23pm        |
| AIR BLK         | .00        | 1:23pm        |
| ACCY CHK        | .08        | 1:24pm        |
| AIR BLK         | .00        | 1:25pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>1:26pm</b> |
| AIR BLK         | .00        | 1:27pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>1:28pm</b> |
| AIR BLK         | .00        | 1:29pm        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*SURRY COUNTY ELKIN PD 850*

Serial Number: 008926      Test Record Number: 840  
Test Date: 02/12/2019      Test Time: 1:30pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:30pm |
| FLO  | Pass   | 1:30pm |
| FC   | Pass   | 1:30pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:31pm |
| SRC  | Pass   | 1:31pm |
| DET  | Pass   | 1:31pm |
| BAR  | Pass   | 1:31pm |
| BT   | Pass   | 1:31pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:31pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:31pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:31pm |
| CAL  | Pass   | 1:31pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008723 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chief R. C. Carter  
Signature of Certifying Official

635  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008723

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734102

Exp Date: 12/07/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 9:55am  |
| AIR BLK  | .00    | 9:56am  |
| ACCY CHK | .08    | 9:57am  |
| AIR BLK  | .00    | 9:58am  |
| SUB TEST | .00    | 9:58am  |
| AIR BLK  | .00    | 9:59am  |
| SUB TEST | .00    | 10:01am |
| AIR BLK  | .00    | 10:02am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008723      Test Record Number: 736  
Test Date: 02/26/2019      Test Time: 10:03am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:03am |
| FLO  | Pass   | 10:03am |
| FC   | Pass   | 10:03am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:03am |
| SRC  | Pass   | 10:03am |
| DET  | Pass   | 10:03am |
| BAR  | Pass   | 10:03am |
| BT   | Pass   | 10:03am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:04am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:04am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:04am |
| CAL  | Pass   | 10:04am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Swain

Instrument Location Swain Co. Jail


Instrument Serial No. 008727

Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul R. Cuth

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008727

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457E

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test.

Lot Number: AG721401


Exp Date: 08/02/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 9:56am  |
| AIR BLK  | .00    | 9:58am  |
| ACCY CHK | .07    | 9:58am  |
| AIR BLK  | .00    | 9:59am  |
| SUB TEST | .00    | 10:00am |
| AIR BLK  | .00    | 10:01am |
| SUB TEST | .00    | 10:02am |
| AIR BLK  | .00    | 10:03am |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

*SWAIN COUNTY SWAIN COUNTY JAIL 860*

Serial Number: 008727      Test Record Number: 1252  
Test Date: 02/26/2019      Test Time: 10:04am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:04am |
| FLO  | Pass   | 10:04am |
| FC   | Pass   | 10:04am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:04am |
| SRC  | Pass   | 10:04am |
| DET  | Pass   | 10:04am |
| BAR  | Pass   | 10:04am |
| BT   | Pass   | 10:04am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:05am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:05am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:05am |
| CAL  | Pass   | 10:05am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Union Instrument Location Union County SO  
Instrument Serial No. 008866 3344 Presson Rd, Monroe

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26th day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hatt

Signature of Certifying Official

650

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008866

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805802


Exp Date: 02/27/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:11pm |
| AIR BLK  | .00    | 12:12pm |
| ACCY CHK | .07    | 12:13pm |
| AIR BLK  | .00    | 12:14pm |
| SUB TEST | .00    | 12:15pm |
| AIR BLK  | .00    | 12:16pm |
| SUB TEST | .00    | 12:17pm |
| AIR BLK  | .00    | 12:18pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008866      Test Record Number: 3125  
Test Date: 02/26/2019      Test Time: 12:19pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:20pm |
| FLO  | Pass   | 12:20pm |
| FC   | Pass   | 12:20pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:20pm |
| SRC  | Pass   | 12:20pm |
| DET  | Pass   | 12:20pm |
| BAR  | Pass   | 12:20pm |
| BT   | Pass   | 12:20pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:21pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:21pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:21pm |
| CAL  | Pass   | 12:21pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Union Instrument Location Union County SO  
Instrument Serial No. 008876 3344 Presson Rd, Monroe

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26th day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hatch  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008876

Test Date: 02/26/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:10pm |
| AIR BLK  | .00    | 12:11pm |
| ACCY CHK | .07    | 12:12pm |
| AIR BLK  | .00    | 12:13pm |
| SUB TEST | .00    | 12:14pm |
| AIR BLK  | .00    | 12:15pm |
| SUB TEST | .00    | 12:16pm |
| AIR BLK  | .00    | 12:17pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008876      Test Record Number: 5224  
Test Date: 02/26/2019      Test Time: 12:19pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:19pm |
| FLO  | Pass   | 12:19pm |
| FC   | Pass   | 12:19pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:19pm |
| SRC  | Pass   | 12:19pm |
| DET  | Pass   | 12:19pm |
| BAR  | Pass   | 12:19pm |
| BT   | Pass   | 12:19pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:20pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:20pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:20pm |
| CAL  | Pass   | 12:20pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Warren Instrument Location Warren Co LEC  
Instrument Serial No. 008793 128 Rafter's Ln, Warrenton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of February, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Barry  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793  
Test Date: 02/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 10:13am |
| AIR BLK  | .00    | 10:14am |
| ACCY CHK | .07    | 10:15am |
| AIR BLK  | .00    | 10:15am |
| SUB TEST | .00    | 10:16am |
| AIR BLK  | .00    | 10:17am |
| SUB TEST | .00    | 10:18am |
| AIR BLK  | .00    | 10:19am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793      Test Record Number: 1406  
Test Date: 02/04/2019      Test Time: 10:20am EST

System Check: Passed

Baseline Tests

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 10:20am |
| FLO  | Pass   | 10:20am |
| FC   | Pass   | 10:20am |

Temperature Tests

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 10:20am |
| SRC  | Pass   | 10:20am |
| DET  | Pass   | 10:20am |
| BAR  | Pass   | 10:20am |
| BT   | Pass   | 10:20am |

Blank Tests

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 10:21am |


Printer Tests

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 10:21am |

CRC Tests

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 10:21am |
| CAL  | Pass   | 10:21am |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne Instrument Location Wayne Co. Detention Ctr.  
Instrument Serial No. 008879 207 E. Chestnut St., Folsom, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879  
Test Date: 02/05/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:30pm |
| AIR BLK  | .00    | 12:31pm |
| ACCY CHK | .07    | 12:32pm |
| AIR BLK  | .00    | 12:33pm |
| SUB TEST | .00    | 12:35pm |
| AIR BLK  | .00    | 12:36pm |
| SUB TEST | .00    | 12:37pm |
| AIR BLK  | .00    | 12:38pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879      Test Record Number: 1157  
Test Date: 02/05/2019      Test Time: 12:38pm EST

System Check: Passed

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:39pm |
| FLO  | Pass   | 12:39pm |
| FC   | Pass   | 12:39pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:39pm |
| SRC  | Pass   | 12:39pm |
| DET  | Pass   | 12:39pm |
| BAR  | Pass   | 12:39pm |
| BT   | Pass   | 12:39pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:40pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:40pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:40pm |
| CAL  | Pass   | 12:40pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wayne Instrument Location Seymour Johnson A.F.B.

Instrument Serial No. 008786 1010 Vermont Garrison Rd., Goldsboro NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21<sup>ST</sup> day of FEBRUARY, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Kiser  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786

Test Date: 02/21/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:06pm |
| AIR BLK  | .00    | 12:07pm |
| ACCY CHK | .08    | 12:08pm |
| AIR BLK  | .00    | 12:09pm |
| SUB TEST | .00    | 12:09pm |
| AIR BLK  | .00    | 12:10pm |
| SUB TEST | .00    | 12:12pm |
| AIR BLK  | .00    | 12:13pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786      Test Record Number: 316  
Test Date: 02/21/2019      Test Time: 12:14pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:14pm |
| FLO  | Pass   | 12:14pm |
| FC   | Pass   | 12:14pm |

Temperature Tests

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:15pm |
| SRC  | Pass   | 12:15pm |
| DET  | Pass   | 12:15pm |
| BAR  | Pass   | 12:15pm |
| BT   | Pass   | 12:15pm |

Blank Tests

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:15pm |


Printer Tests

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:15pm |

CRC Tests

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:15pm |
| CAL  | Pass   | 12:15pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne


Instrument Location Wayne Co. Detention Ctr.

Instrument Serial No. 008508 207 E CHESTNUT ST, GOLDSBORO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of FEBRUARY, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Keelsee  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008588

Test Date: 02/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:43am |
| AIR BLK  | .00    | 11:44am |
| ACCY CHK | .08    | 11:44am |
| AIR BLK  | .00    | 11:45am |
| SUB TEST | .00    | 11:46am |
| AIR BLK  | .00    | 11:46am |
| SUB TEST | .00    | 11:48am |
| AIR BLK  | .00    | 11:49am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008588      Test Record Number: 1001  
Test Date: 02/01/2019      Test Time: 11:50am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:50am |
| FLO  | Pass   | 11:50am |
| FC   | Pass   | 11:50am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:50am |
| SRC  | Pass   | 11:50am |
| DET  | Pass   | 11:50am |
| BAR  | Pass   | 11:50am |
| BT   | Pass   | 11:50am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:51am |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:51am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:51am |
| CAL  | Pass   | 11:51am |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wayne Instrument Location Wayne Co. Detention Ctr.

Instrument Serial No. 008671 207 E. CHESTNUT ST., GOLDSBORO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28<sup>th</sup> day of FEBRUARY, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Kiser  
Signature of Certifying Official

687  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671

Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 12:46pm |
| AIR BLK  | .00    | 12:47pm |
| ACCY CHK | .08    | 12:48pm |
| AIR BLK  | .00    | 12:49pm |
| SUB TEST | .00    | 12:49pm |
| AIR BLK  | .00    | 12:50pm |
| SUB TEST | .00    | 12:51pm |
| AIR BLK  | .00    | 12:52pm |

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671      Test Record Number: 4749  
Test Date: 02/28/2019      Test Time: 12:53pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 12:53pm |
| FLO  | Pass   | 12:53pm |
| FC   | Pass   | 12:54pm |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 12:54pm |
| SRC  | Pass   | 12:54pm |
| DET  | Pass   | 12:54pm |
| BAR  | Pass   | 12:54pm |
| BT   | Pass   | 12:54pm |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 12:54pm |


**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 12:54pm |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 12:55pm |
| CAL  | Pass   | 12:55pm |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne Instrument Location Wayne Co. Detention Ctr.  
Instrument Serial No. 008649 207 E. Chestnut St., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28<sup>th</sup> day of FEBRUARY, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649

Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:01pm |
| AIR BLK  | .00    | 1:02pm |
| ACCY CHK | .07    | 1:03pm |
| AIR BLK  | .00    | 1:04pm |
| SUB TEST | .00    | 1:05pm |
| AIR BLK  | .00    | 1:06pm |
| SUB TEST | .00    | 1:08pm |
| AIR BLK  | .00    | 1:09pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649      Test Record Number: 3733  
Test Date: 02/28/2019      Test Time: 1:11pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:11pm |
| FLO  | Pass   | 1:11pm |
| FC   | Pass   | 1:11pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:11pm |
| SRC  | Pass   | 1:11pm |
| DET  | Pass   | 1:11pm |
| BAR  | Pass   | 1:11pm |
| BT   | Pass   | 1:11pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:12pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:12pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:12pm |
| CAL  | Pass   | 1:12pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne Instrument Location Wayne Co. Detention CN.  
Instrument Serial No. 008879 207 E. Chestnut St., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of FEBRUARY, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



James A. Keese  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879

Test Date: 02/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 1:02pm |
| AIR BLK  | .00    | 1:03pm |
| ACCY CHK | .07    | 1:04pm |
| AIR BLK  | .00    | 1:05pm |
| SUB TEST | .00    | 1:06pm |
| AIR BLK  | .00    | 1:07pm |
| SUB TEST | .00    | 1:09pm |
| AIR BLK  | .00    | 1:10pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879      Test Record Number: 1165  
Test Date: 02/28/2019      Test Time: 1:11pm EST

System Check: Passed

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 1:12pm |
| FLO  | Pass   | 1:12pm |
| FC   | Pass   | 1:12pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 1:12pm |
| SRC  | Pass   | 1:12pm |
| DET  | Pass   | 1:12pm |
| BAR  | Pass   | 1:12pm |
| BT   | Pass   | 1:12pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 1:12pm |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 1:13pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 1:13pm |
| CAL  | Pass   | 1:13pm |

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Watauga Instrument Location Boone PD  
Instrument Serial No. 008716 Boone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WATAUGA COUNTY BOONE P D 940

Serial Number: 008716  
Test Date: 02/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

| Test            | g/210L     | Time          |
|-----------------|------------|---------------|
| DIAG            | Pass       | 3:33pm        |
| AIR BLK         | .00        | 3:34pm        |
| ACCY CHK        | .08        | 3:35pm        |
| AIR BLK         | .00        | 3:36pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>3:36pm</b> |
| AIR BLK         | .00        | 3:37pm        |
| <b>SUB TEST</b> | <b>.00</b> | <b>3:39pm</b> |
| AIR BLK         | .00        | 3:40pm        |

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WATAUGA COUNTY BOONE P D 940

Serial Number: 008716      Test Record Number: 2394  
Test Date: 02/25/2019      Test Time: 3:41pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 3:41pm |
| FLO  | Pass   | 3:41pm |
| FC   | Pass   | 3:42pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 3:42pm |
| SRC  | Pass   | 3:42pm |
| DET  | Pass   | 3:42pm |
| BAR  | Pass   | 3:42pm |
| BT   | Pass   | 3:42pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 3:42pm |

**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 3:42pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 3:43pm |
| CAL  | Pass   | 3:43pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wilkes Instrument Location Wilkes County  
Instrument Serial No. 008865 Detention Center  
Wilkesboro N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of February, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865

Test Date: 02/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

| Test            | g/210L     | Time           |
|-----------------|------------|----------------|
| DIAG            | Pass       | 11:16am        |
| AIR BLK         | .00        | 11:17am        |
| ACCY CHK        | .07        | 11:17am        |
| AIR BLK         | .00        | 11:19am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:20am</b> |
| AIR BLK         | .00        | 11:21am        |
| <b>SUB TEST</b> | <b>.00</b> | <b>11:22am</b> |
| AIR BLK         | .00        | 11:23am        |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**WILKES COUNTY WILKES CO DETENTION 960**

Serial Number: 008865      Test Record Number: 673  
Test Date: 02/12/2019      Test Time: 11:24am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:24am |
| FLO  | Pass   | 11:24am |
| FC   | Pass   | 11:24am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:24am |
| SRC  | Pass   | 11:24am |
| DET  | Pass   | 11:24am |
| BAR  | Pass   | 11:24am |
| BT   | Pass   | 11:24am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:25am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:25am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:25am |
| CAL  | Pass   | 11:25am |

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County

Wilkes

Instrument Location

Wilkes County

Instrument Serial No.

008843

Detention Center  
Wilkesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843  
Test Date: 02/12/2019

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

| Test     | g/210L | Time    |
|----------|--------|---------|
| DIAG     | Pass   | 11:03am |
| AIR BLK  | .00    | 11:04am |
| ACCY CHK | .08    | 11:04am |
| AIR BLK  | .00    | 11:05am |
| SUB TEST | .00    | 11:06am |
| AIR BLK  | .00    | 11:07am |
| SUB TEST | .00    | 11:09am |
| AIR BLK  | .00    | 11:10am |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**WILKES COUNTY WILKES CO DETENTION 960**

Serial Number: 008843      Test Record Number: 2313  
Test Date: 02/12/2019      Test Time: 11:11am EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time    |
|------|--------|---------|
| IR   | Pass   | 11:11am |
| FLO  | Pass   | 11:11am |
| FC   | Pass   | 11:11am |

**Temperature Tests**

| Test | Status | Time    |
|------|--------|---------|
| FC1  | Pass   | 11:11am |
| SRC  | Pass   | 11:11am |
| DET  | Pass   | 11:11am |
| BAR  | Pass   | 11:11am |
| BT   | Pass   | 11:11am |

**Blank Tests**

| Test | Status | Time    |
|------|--------|---------|
| AIR  | Pass   | 11:12am |

**Printer Tests**

| Test | Status | Time    |
|------|--------|---------|
| PRNT | Pass   | 11:12am |

**CRC Tests**

| Test | Status | Time    |
|------|--------|---------|
| COMP | Pass   | 11:12am |
| CAL  | Pass   | 11:12am |

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County

Yadkin

Instrument Location

Yadkin County

Instrument Serial No.

008854


Jail

Yadkinville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



C. Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008854

Test Date: 02/19/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821801


Exp Date: 08/06/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:24pm |
| AIR BLK  | .00    | 2:25pm |
| ACCY CHK | .08    | 2:26pm |
| AIR BLK  | .00    | 2:27pm |
| SUB TEST | .00    | 2:28pm |
| AIR BLK  | .00    | 2:29pm |
| SUB TEST | .00    | 2:30pm |
| AIR BLK  | .00    | 2:31pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**YADKIN COUNTY YADKIN CO JAIL 980**

Serial Number: 008854      Test Record Number: 572  
Test Date: 02/19/2019      Test Time: 2:33pm EST

System Check: *Passed*

**Baseline Tests**

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:33pm |
| FLO  | Pass   | 2:33pm |
| FC   | Pass   | 2:33pm |

**Temperature Tests**

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:33pm |
| SRC  | Pass   | 2:33pm |
| DET  | Pass   | 2:33pm |
| BAR  | Pass   | 2:33pm |
| BT   | Pass   | 2:33pm |

**Blank Tests**

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:34pm |


**Printer Tests**

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:34pm |

**CRC Tests**

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:34pm |
| CAL  | Pass   | 2:34pm |

Preventive Maintenance  
Status: Pass

  
**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Yadkin Instrument Location Yadkin County  
Instrument Serial No. 008944 Jail  
Yadkinville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19<sup>th</sup> day of February, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



C. J. Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944

Test Date: 02/19/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

| Test     | g/210L | Time   |
|----------|--------|--------|
| DIAG     | Pass   | 2:12pm |
| AIR BLK  | .00    | 2:13pm |
| ACCY CHK | .07    | 2:13pm |
| AIR BLK  | .00    | 2:15pm |
| SUB TEST | .00    | 2:16pm |
| AIR BLK  | .00    | 2:17pm |
| SUB TEST | .00    | 2:18pm |
| AIR BLK  | .00    | 2:19pm |

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944      Test Record Number: 1550  
Test Date: 02/19/2019      Test Time: 2:20pm EST

System Check: *Passed*

Baseline Tests

| Test | Status | Time   |
|------|--------|--------|
| IR   | Pass   | 2:20pm |
| FLO  | Pass   | 2:20pm |
| FC   | Pass   | 2:21pm |

Temperature Tests

| Test | Status | Time   |
|------|--------|--------|
| FC1  | Pass   | 2:21pm |
| SRC  | Pass   | 2:21pm |
| DET  | Pass   | 2:21pm |
| BAR  | Pass   | 2:21pm |
| BT   | Pass   | 2:21pm |

Blank Tests

| Test | Status | Time   |
|------|--------|--------|
| AIR  | Pass   | 2:21pm |


Printer Tests

| Test | Status | Time   |
|------|--------|--------|
| PRNT | Pass   | 2:21pm |

CRC Tests

| Test | Status | Time   |
|------|--------|--------|
| COMP | Pass   | 2:22pm |
| CAL  | Pass   | 2:22pm |

Preventive Maintenance  
Status: Pass

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007