

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location ALAMANCE CO JAIL

Instrument Serial No. 008924 109 S. Maple ST
GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of February, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUTY ALAMANCE CO. JAIL 000

Serial Number: 008924

Test Date: 02/17/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	3:51pm
AIR BLK	.00	3:52pm
ACCY CHK	.08	3:53pm
AIR BLK	.00	3:54pm
SUB TEST	.00	3:54pm
AIR BLK	.00	3:55pm
SUB TEST	.00	3:57pm
AIR BLK	.00	3:58pm

Reported/AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUTY ALAMANCE CO. JAIL 000

Serial Number: 008924 Test Record Number: 1358
Test Date: 02/17/2018 Test Time: 3:59pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:59pm
FLO	Pass	3:59pm
FC	Pass	3:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:59pm
SRC	Pass	3:59pm
DET	Pass	3:59pm
BAR	Pass	3:59pm
BT	Pass	3:59pm

Blank Tests

Test	Status	Time
AIR	Pass	4:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:00pm

CRC Tests

Test	Status	Time
COMP	Pass	4:00pm
CAL	Pass	4:00pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Alexander Instrument Location Alexander County S.D.
Instrument Serial No. 008881 91 Commercial Park Ave., Taybesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALEXANDER COUNTY ALEXANDER COUNTY SD
010

Serial Number: 008881

Test Date: 02/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019


Test	g/210L	Time
DIAG	Pass	9:11am
AIR BLK	.00	9:12am
ACCY CHK	.07	9:12am
AIR BLK	.00	9:13am
SUB TEST	.00	9:14am
AIR BLK	.00	9:15am
SUB TEST	.00	9:16am
AIR BLK	.00	9:18am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

ALEXANDER COUNTY ALEXANDER COUNTY SD 010

Serial Number: 008881 Test Record Number: 659
Test Date: 02/06/2018 Test Time: 9:19am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:19am
FLO	Pass	9:19am
FC	Pass	9:19am

Temperature Tests

Test	Status	Time
FC1	Pass	9:20am
SRC	Pass	9:20am
DET	Pass	9:20am
BAR	Pass	9:20am
BT	Pass	9:20am

Blank Tests

Test	Status	Time
AIR	Pass	9:20am


Printer Tests

Test	Status	Time
PRNT	Pass	9:20am

CRC Tests

Test	Status	Time
COMP	Pass	9:20am
CAL	Pass	9:20am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Alexander Instrument Location Alexander County SD
Instrument Serial No. 008694 91 Commercial Park Ave., Taylorsville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALEXANDER COUNTY ALEXANDER COUNTY SD
010

Serial Number: 008694
Test Date: 02/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test g/210L Time

DIAG	Pass	8:50am
AIR BLK	.00	8:51am
ACCY CHK	.07	8:51am
AIR BLK	.00	8:52am
SUB TEST	.00	8:53am
AIR BLK	.00	8:54am
SUB TEST	.00	8:55am
AIR BLK	.00	8:56am

Reported AC: 00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALEXANDER COUNTY ALEXANDER COUNTY SD 010

Serial Number: 008694 Test Record Number: 1325
Test Date: 02/15/2018 Test Time: 8:58am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	8:58am
FLO	Pass	8:58am
FC	Pass	8:59am

Temperature Tests

Test	Status	Time
FC1	Pass	8:59am
SRC	Pass	8:59am
DET	Pass	8:59am
BAR	Pass	8:59am
BT	Pass	8:59am

Blank Tests

Test	Status	Time
AIR	Pass	8:59am


Printer Tests

Test	Status	Time
PRNT	Pass	8:59am

CRC Tests

Test	Status	Time
COMP	Pass	8:59am
CAL	Pass	8:59am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County

ANSON Co.

Instrument Location

ANSON Co. Sheriff's Office

Instrument Serial No.


008597

Wadesboro NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Signature of Certifying Official

Certificate Number

654

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ANSON COUNTY ANSON COUNTY SO. 030

Serial Number: 008597

Test Date: 02/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test g/210L Time

DIAG Pass 6:09pm

AIR BLK .00 6:10pm

ACCY CHK .07 6:11pm

AIR BLK .00 6:12pm


SUB TEST .00 6:13pm

AIR BLK .00 6:13pm


SUB TEST .00 6:15pm

AIR BLK .00 6:16pm

Reported At: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY ANSON COUNTY SO. 030

Serial Number: 008597 Test Record Number: 1666
Test Date: 02/06/2018 Test Time: 6:18pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:19pm
FLO	Pass	6:19pm
FC	Pass	6:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:19pm
SRC	Pass	6:19pm
DET	Pass	6:19pm
BAR	Pass	6:19pm
BT	Pass	6:19pm

Blank Tests

Test	Status	Time
AIR	Pass	6:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	6:19pm

CRC Tests

Test	Status	Time
COMP	Pass	6:20pm
CAL	Pass	6:20pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ANSON Co. Instrument Location ANSON Co. Sheriff Office
Instrument Serial No. 008739 Wadesboro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ANSON COUNTY ANSON COUNTY S.O. 030

Serial Number: 008739

Test Date: 02/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C.

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test g/210L Time

DIAG Pass 6:11pm

AIR BLK .00 6:12pm

ACCY CHK .07 6:13pm

AIR BLK .00 6:14pm


SUB TEST .00 6:14pm

AIR BLK .00 6:15pm


SUB TEST .00 6:17pm

AIR BLK .00 6:18pm

Reported AG: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY ANSON COUNTY S.O. 030

Serial Number: 008739 Test Record Number: 257
Test Date: 02/06/2018 Test Time: 6:19pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:19pm
FLO	Pass	6:19pm
FC	Pass	6:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:19pm
SRC	Pass	6:19pm
DET	Pass	6:19pm
BAR	Pass	6:19pm
BT	Pass	6:19pm

Blank Tests

Test	Status	Time
AIR	Pass	6:20pm


Printer Tests

Test	Status	Time
PRNT	Pass	6:20pm

CRC Tests

Test	Status	Time
COMP	Pass	6:20pm
CAL	Pass	6:20pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Beaufort Instrument Location Belhaven P.D.
Instrument Serial No. 008928 Belhaven, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. O.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: 008928

Test Date: 02/21/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	10:39am
AIR BLK	.00	10:40am
ACCY CHK	.08	10:41am
AIR BLK	.00	10:42am
SUB TEST	.00	10:42am
AIR BLK	.00	10:43am
SUB TEST	.00	10:45am
AIR BLK	.00	10:46am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: 008928 Test Record Number: 336
Test Date: 02/21/2018 Test Time: 10:47am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:47am
FLO	Pass	10:47am
FC	Pass	10:47am

Temperature Tests

Test	Status	Time
FC1	Pass	10:47am
SRC	Pass	10:47am
DET	Pass	10:47am
BAR	Pass	10:47am
BT	Pass	10:47am

Blank Tests

Test	Status	Time
AIR	Pass	10:48am

Printer Tests

Test	Status	Time
PRNT	Pass	10:48am

CRC Tests

Test	Status	Time
COMP	Pass	10:48am
CAL	Pass	10:48am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Bertie Instrument Location Bertie Co. S.O.
Instrument Serial No. 008897 222 County Farm Rd., Windsor, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key M. P.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BERTIE COUNTY BERTIE CO SO 070

Serial Number: 008897

Test Date: 02/20/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019


Test	g/210L	Time
DIAG	Pass	10:10am
AIR BLK	.00	10:11am
ACCY CHK	.08	10:12am
AIR BLK	.00	10:13am
SUB TEST	.00	10:14am
AIR BLK	.00	10:15am
SUB TEST	.00	10:16am
AIR BLK	.00	10:17am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BERTIE COUNTY BERTIE CO SO 070

Serial Number: 008897 Test Record Number: 1106
Test Date: 02/20/2018 Test Time: 10:18am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:18am
FLO	Pass	10:18am
FC	Pass	10:18am

Temperature Tests

Test	Status	Time
FC1	Pass	10:18am
SRC	Pass	10:18am
DET	Pass	10:18am
BAR	Pass	10:18am
BT	Pass	10:18am

Blank Tests

Test	Status	Time
AIR	Pass	10:19am


Printer Tests

Test	Status	Time
PRNT	Pass	10:19am

CRC Tests

Test	Status	Time
COMP	Pass	10:19am
CAL	Pass	10:19am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Buncombe Instrument Location Buncombe Co. Jail
Instrument Serial No. 008697 Asheville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL
100

Serial Number: 008697
Test Date: 02/28/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716202

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	1:04pm
AIR BLK	.00	1:05pm
ACCY CHK	.08	1:06pm
AIR BLK	.00	1:07pm
SUB TEST	.00	1:07pm
AIR BLK	.00	1:08pm
SUB TEST	.00	1:10pm
AIR BLK	.00	1:11pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL 100

Serial Number: 008697 Test Record Number: 3326
Test Date: 02/28/2018 Test Time: 1:17pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:17pm
FLO	Pass	1:17pm
FC	Pass	1:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:18pm
SRC	Pass	1:18pm
DET	Pass	1:18pm
BAR	Pass	1:18pm
BT	Pass	1:18pm

Blank Tests

Test	Status	Time
AIR	Pass	1:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:18pm

CRC Tests

Test	Status	Time
COMP	Pass	1:19pm
CAL	Pass	1:19pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Buncombe Instrument Location Buncombe Co. Jail

Instrument Serial No. 008431 Asheville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL
100

Serial Number: 008631
Test Date: 02/28/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716202

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	12:31pm
AIR BLK	.00	12:32pm
ACCY CHK	.07	12:33pm
AIR BLK	.00	12:34pm
SUB TEST	.00	12:35pm
AIR BLK	.00	12:36pm
SUB TEST	.00	12:37pm
AIR BLK	.00	12:38pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL 100

Serial Number: 008631 Test Record Number: 4966
Test Date: 02/28/2018 Test Time: 12:39pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:39pm
FLO	Pass	12:39pm
FC	Pass	12:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:39pm
SRC	Pass	12:39pm
DET	Pass	12:39pm
BAR	Pass	12:39pm
BT	Pass	12:39pm

Blank Tests

Test	Status	Time
AIR	Pass	12:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:40pm

CRC Tests

Test	Status	Time
COMP	Pass	12:40pm
CAL	Pass	12:40pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location Catawba County SD
Instrument Serial No. 0088821 100 B Southwest Blvd, Newton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821
Test Date: 02/21/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	3:46pm
AIR BLK	.00	3:47pm
ACCY CHK	.07	3:48pm
AIR BLK	.00	3:48pm
SUB TEST	.00	3:49pm
AIR BLK	.00	3:50pm
SUB TEST	.00	3:52pm
AIR BLK	.00	3:53pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821 Test Record Number: 1872
Test Date: 02/21/2018 Test Time: 3:54pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:54pm
FLO	Pass	3:54pm
FC	Pass	3:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:54pm
SRC	Pass	3:54pm
DET	Pass	3:54pm
BAR	Pass	3:54pm
BT	Pass	3:54pm

Blank Tests

Test	Status	Time
AIR	Pass	3:55pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:55pm

CRC Tests

Test	Status	Time
COMP	Pass	3:55pm
CAL	Pass	3:55pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chatham Co. Instrument Location Siler City Police Dept.
Instrument Serial No. 008811 Siler City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHATHAM COUNTY SILER CITY PD. 180

Serial Number: 008811

Test Date: 02/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	4:42pm
AIR BLK	.00	4:43pm
ACCY CHK	.07	4:44pm
AIR BLK	.00	4:45pm
SUB TEST	.00	4:46pm
AIR BLK	.00	4:47pm
SUB TEST	.00	4:48pm
AIR BLK	.00	4:49pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CHATHAM COUNTY SILER CITY PD. 180

Serial Number: 008811 Test Record Number: 1279

Test Date: 02/14/2018 Test Time: 4:49pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	4:50pm
FLO	Pass	4:50pm
FC	Pass	4:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:50pm
SRC	Pass	4:50pm
DET	Pass	4:50pm
BAR	Pass	4:50pm
BT	Pass	4:50pm

Blank Tests

Test	Status	Time
AIR	Pass	4:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:51pm

CRC Tests

Test	Status	Time
COMP	Pass	4:51pm
CAL	Pass	4:51pm

**Preventive Maintenance
Status: Pass**


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN

Instrument Location BAT MOBILE UNIT 12

Instrument Serial No. 008600

HAVELOCK, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Basso
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 12 240

Serial Number: 008600

Test Date: 02/10/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	11:14pm
AIR BLK	.00	11:15pm
ACCY CHK	.07	11:15pm
AIR BLK	.00	11:16pm
SUB TEST	.00	11:17pm
AIR BLK	.00	11:18pm
SUB TEST	.00	11:19pm
AIR BLK	.00	11:20pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 12 240

Serial Number: 008600 Test Record Number: 1789
Test Date: 02/10/2018 Test Time: 11:22pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:22pm
FLO	Pass	11:22pm
FC	Pass	11:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:22pm
SRC	Pass	11:22pm
DET	Pass	11:22pm
BAR	Pass	11:22pm
BT	Pass	11:22pm

Blank Tests

Test	Status	Time
AIR	Pass	11:23pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:23pm

CRC Tests

Test	Status	Time
COMP	Pass	11:23pm
CAL	Pass	11:23pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Currituck Instrument Location Currituck Co. S.O.
Instrument Serial No. 008947 407-A Maple Rd., Maple, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27th day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Tina A. Heese
Signature of Certifying Official

447
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CURRITUCK COUNTY CURRITUCK SO-MAPLE 260

Serial Number: 008947 Test Record Number: 2153
Test Date: 02/27/2018 Test Time: 10:54am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:54am
FLO	Pass	10:54am
FC	Pass	10:54am

Temperature Tests

Test	Status	Time
FC1	Pass	10:54am
SRC	Pass	10:54am
DET	Pass	10:54am
BAR	Pass	10:54am
BT	Pass	10:54am

Blank Tests

Test	Status	Time
AIR	Pass	10:55am


Printer Tests

Test	Status	Time
PRNT	Pass	10:55am

CRC Tests

Test	Status	Time
COMP	Pass	10:56am
CAL	Pass	10:56am

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

CURRITUCK COUNTY CURRITUCK SO-MAPLE
260

Serial Number: 008947
Test Date: 02/27/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302


Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	10:58am
AIR BLK	.00	10:59am
ACCY CHK	.07	10:59am
AIR BLK	.00	11:00am
SUB TEST	.00	11:01am
AIR BLK	.00	11:02am
SUB TEST	.00	11:03am
AIR BLK	.00	11:04am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County DAVIDSON Instrument Location DAVIDSON CO. JAIL

Instrument Serial No. 008845 Lexington, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of FEBRUARY, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. H. Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845

Test Date: 02/27/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	2:01pm
AIR BLK	.00	2:02pm
ACCY CHK	.08	2:02pm
AIR BLK	.00	2:04pm
SUB TEST	.00	2:04pm
AIR BLK	.00	2:05pm
SUB TEST	.00	2:07pm
AIR BLK	.00	2:08pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845 Test Record Number: 2736

Test Date: 02/27/2018 Test Time: 2:10pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:11pm
FLO	Pass	2:11pm
FC	Pass	2:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:11pm
SRC	Pass	2:11pm
DET	Pass	2:11pm
BAR	Pass	2:11pm
BT	Pass	2:11pm

Blank Tests

Test	Status	Time
AIR	Pass	2:11pm

Printer Tests


Test	Status	Time
PRNT	Pass	2:12pm

CRC Tests

Test	Status	Time
COMP	Pass	2:12pm
CAL	Pass	2:12pm

Preventive Maintenance

Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County

Davie

Instrument Location

Davie County Jail

Instrument Serial No.


008905

Mocksville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

657

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905

Test Date: 02/12/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401


Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	11:22am
AIR BLK	.00	11:22am
ACCY CHK	.08	11:23am
AIR BLK	.00	11:24am
SUB TEST	.00	11:25am
AIR BLK	.00	11:26am
SUB TEST	.00	11:27am
AIR BLK	.00	11:28am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905 Test Record Number: 2157
Test Date: 02/12/2018 Test Time: 11:29am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:29am
FLO	Pass	11:29am
FC	Pass	11:29am

Temperature Tests

Test	Status	Time
FC1	Pass	11:29am
SRC	Pass	11:29am
DET	Pass	11:29am
BAR	Pass	11:29am
BT	Pass	11:29am

Blank Tests

Test	Status	Time
AIR	Pass	11:30am


Printer Tests

Test	Status	Time
PRNT	Pass	11:30am

CRC Tests

Test	Status	Time
COMP	Pass	11:30am
CAL	Pass	11:30am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Edgecombe Instrument Location Edgecombe Co. Magistrate's

Instrument Serial No. 008663 Office, 3005. Anaconda Rd., Tarboro, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR
320

Serial Number: 008663
Test Date: 02/23/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	9:41am
AIR BLK	.00	9:42am
ACCY CHK	.08	9:43am
AIR BLK	.00	9:44am
SUB TEST	.00	9:45am
AIR BLK	.00	9:45am
SUB TEST	.00	9:47am
AIR BLK	.00	9:48am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR 320

Serial Number: 008663 Test Record Number: 2836
Test Date: 02/23/2018 Test Time: 9:54am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:55am
FLO	Pass	9:55am
FC	Pass	9:55am

Temperature Tests

Test	Status	Time
FC1	Pass	9:55am
SRC	Pass	9:55am
DET	Pass	9:55am
BAR	Pass	9:55am
BT	Pass	9:55am

Blank Tests

Test	Status	Time
AIR	Pass	9:56am

Printer Tests

Test	Status	Time
PRNT	Pass	9:56am

CRC Tests

Test	Status	Time
COMP	Pass	9:56am
CAL	Pass	9:56am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Edgercombe Instrument Location Edgercombe Co. Magistrate's
Instrument Serial No. 008603 Office, 300 S. Anaconda Rd., Tarboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kay S. O.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR
320

Serial Number: 008603
Test Date: 02/23/2018

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702402

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	9:19am
AIR BLK	.00	9:20am
ACCY CHK	.08	9:21am
AIR BLK	.00	9:22am
SUB TEST	.00	9:23am
AIR BLK	.00	9:24am
SUB TEST	.00	9:25am
AIR BLK	.00	9:26am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR 320

Serial Number: 008603 Test Record Number: 1652
Test Date: 02/23/2018 Test Time: 9:34am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:34am
FLO	Pass	9:34am
FC	Pass	9:34am

Temperature Tests

Test	Status	Time
FC1	Pass	9:34am
SRC	Pass	9:34am
DET	Pass	9:34am
BAR	Pass	9:34am
BT	Pass	9:34am

Blank Tests

Test	Status	Time
AIR	Pass	9:35am

Printer Tests

Test	Status	Time
PRNT	Pass	9:35am

CRC Tests

Test	Status	Time
COMP	Pass	9:35am
CAL	Pass	9:35am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County FORSYTH Instrument Location KERNERSVILLE Police
Instrument Serial No. 008650 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650
Test Date: 02/06/2018

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	3:33pm
AIR BLK	.00	3:33pm
ACCY CHK	.08	3:34pm
AIR BLK	.00	3:35pm
SUB TEST	.00	3:36pm
AIR BLK	.00	3:37pm
SUB TEST	.00	3:38pm
AIR BLK	.00	3:39pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650 Test Record Number: 1395

Test Date: 02/06/2018 Test Time: 3:40pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:41pm
FLO	Pass	3:41pm
FC	Pass	3:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:41pm
SRC	Pass	3:41pm
DET	Pass	3:41pm
BAR	Pass	3:41pm
BT	Pass	3:41pm

Blank Tests

Test	Status	Time
AIR	Pass	3:41pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:42pm

CRC Tests

Test	Status	Time
COMP	Pass	3:42pm
CAL	Pass	3:42pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention
Instrument Serial No. 008925 Winston-Salem, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kurt H. [Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: 008925
Test Date: 02/06/2018

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	4:27pm
AIR BLK	.00	4:28pm
ACCY CHK	.07	4:28pm
AIR BLK	.00	4:30pm
SUB TEST	.00	4:31pm
AIR BLK	.00	4:32pm
SUB TEST	.00	4:34pm
AIR BLK	.00	4:35pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008925 Test Record Number: 2093
Test Date: 02/06/2018 Test Time: 4:36pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:36pm
FLO	Pass	4:36pm
FC	Pass	4:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:36pm
SRC	Pass	4:36pm
DET	Pass	4:36pm
BAR	Pass	4:36pm
BT	Pass	4:36pm

Blank Tests

Test	Status	Time
AIR	Pass	4:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:37pm

CRC Tests

Test	Status	Time
COMP	Pass	4:37pm
CAL	Pass	4:37pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention
Instrument Serial No. 008583 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION

330

Serial Number: 008583

Test Date: 02/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	4:26pm
AIR BLK	.00	4:27pm
ACCY CHK	.08	4:27pm
AIR BLK	.00	4:29pm
SUB TEST	.00	4:29pm
AIR BLK	.00	4:30pm
SUB TEST	.00	4:32pm
AIR BLK	.00	4:33pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008583 Test Record Number: 7069
Test Date: 02/06/2018 Test Time: 4:34pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:34pm
FLO	Pass	4:34pm
FC	Pass	4:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:34pm
SRC	Pass	4:34pm
DET	Pass	4:34pm
BAR	Pass	4:34pm
BT	Pass	4:34pm

Blank Tests

Test	Status	Time
AIR	Pass	4:35pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:35pm

CRC Tests

Test	Status	Time
COMP	Pass	4:35pm
CAL	Pass	4:35pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention
Instrument Serial No. 008659 Winston-Salem, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6 day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION

330

Serial Number: 008659

Test Date: 02/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test g/210L Time

DIAG Pass 4:25pm

AIR BLK .00 4:26pm

ACCY CHK .07 4:26pm

AIR BLK .00 4:27pm


SUB TEST .00 4:28pm

AIR BLK .00 4:29pm


SUB TEST .00 4:31pm

AIR BLK .00 4:32pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008659 Test Record Number: 4653
Test Date: 02/06/2018 Test Time: 4:33pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:33pm
FLO	Pass	4:33pm
FC	Pass	4:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:33pm
SRC	Pass	4:33pm
DET	Pass	4:33pm
BAR	Pass	4:33pm
BT	Pass	4:33pm

Blank Tests

Test	Status	Time
AIR	Pass	4:34pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:34pm

CRC Tests

Test	Status	Time
COMP	Pass	4:34pm
CAL	Pass	4:34pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County Gaston Instrument Location But mobile Unit 11

Instrument Serial No. 008770

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of February, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008970

Test Date: 02/03/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test


Lot Number: AG607501

Exp Date: 03/15/2018

Test g/210L Time


DIAG	Pass	9:14pm
AIR BLK	.00	9:15pm
ACCY CHK	.07	9:16pm
AIR BLK	.00	9:16pm
SUB TEST	.00	9:17pm
AIR BLK	.00	9:18pm
SUB TEST	.00	9:19pm
AIR BLK	.00	9:20pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008970 Test Record Number: 433
Test Date: 02/03/2018 Test Time: 9:21pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:22pm
FLO	Pass	9:22pm
FC	Pass	9:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:22pm
SRC	Pass	9:22pm
DET	Pass	9:22pm
BAR	Pass	9:22pm
BT	Pass	9:22pm

Blank Tests

Test	Status	Time
AIR	Pass	9:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:23pm

CRC Tests

Test	Status	Time
COMP	Pass	9:23pm
CAL	Pass	9:23pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Gaston Instrument Location Bat mobile Unit 11

Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of February, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris Jones
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008973

Test Date: 02/03/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501


Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	9:13pm
AIR BLK	.00	9:14pm
ACCY CHK	.08	9:15pm
AIR BLK	.00	9:16pm
SUB TEST	.00	9:16pm
AIR BLK	.00	9:18pm
SUB TEST	.00	9:19pm
AIR BLK	.00	9:20pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008973 Test Record Number: 454
Test Date: 02/03/2018 Test Time: 9:21pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:21pm
FLO	Pass	9:21pm
FC	Pass	9:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:21pm
SRC	Pass	9:21pm
DET	Pass	9:21pm
BAR	Pass	9:21pm
BT	Pass	9:21pm

Blank Tests

Test	Status	Time
AIR	Pass	9:22pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:22pm

CRC Tests

Test	Status	Time
COMP	Pass	9:22pm
CAL	Pass	9:22pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gates Instrument Location Gates Co. S.O.

Instrument Serial No. 008884 202 Court St., Gatesville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GATES COUNTY GATES CO SO 360

Serial Number: 008884

Test Date: 02/22/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	1:04pm
AIR BLK	.00	1:05pm
ACCY CHK	.07	1:06pm
AIR BLK	.00	1:07pm
SUB TEST	.00	1:07pm
AIR BLK	.00	1:08pm
SUB TEST	.00	1:10pm
AIR BLK	.00	1:11pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GATES COUNTY GATES CO SO 360

Serial Number: 008884 Test Record Number: 842
Test Date: 02/22/2018 Test Time: 1:12pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:12pm
FLO	Pass	1:12pm
FC	Pass	1:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:12pm
SRC	Pass	1:12pm
DET	Pass	1:12pm
BAR	Pass	1:12pm
BT	Pass	1:12pm

Blank Tests

Test	Status	Time
AIR	Pass	1:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:13pm

CRC Tests

Test	Status	Time
COMP	Pass	1:13pm
CAL	Pass	1:13pm

Preventive Maintenance
Status: Pass

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location High Point
Instrument Serial No. 008896 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of FEBRUARY, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT PD 401

Serial Number: 008896

Test Date: 02/05/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	12:08pm
AIR BLK	.00	12:09pm
ACCY CHK	.07	12:10pm
AIR BLK	.00	12:10pm
SUB TEST	.00	12:12pm
AIR BLK	.00	12:13pm
SUB TEST	.00	12:14pm
AIR BLK	.00	12:15pm

Reported AC: .00 g/210L

L. Kun Dean
Signature of Chemical Analyst

Court CVR

L. Kun Dean
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT PD 401

Serial Number: 008896 Test Record Number: 1062
Test Date: 02/05/2018 Test Time: 12:03pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:04pm
FLO	Pass	12:04pm
FC	Pass	12:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:04pm
SRC	Pass	12:04pm
DET	Pass	12:04pm
BAR	Pass	12:04pm
BT	Pass	12:04pm

Blank Tests

Test	Status	Time
AIR	Pass	12:04pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:05pm

CRC Tests

Test	Status	Time
COMP	Pass	12:05pm
CAL	Pass	12:05pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford


Instrument Location Greensboro PD

Instrument Serial No. 008725 100 Police Plaza Greensboro, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725

Test Date: 02/26/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	3:09pm
AIR BLK	.00	3:09pm
ACCY CHK	.08	3:10pm
AIR BLK	.00	3:12pm
SUB TEST	.00	3:12pm
AIR BLK	.00	3:13pm
SUB TEST	.00	3:15pm
AIR BLK	.00	3:16pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725 Test Record Number: 4114
Test Date: 02/26/2018 Test Time: 3:17pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:18pm
FLO	Pass	3:18pm
FC	Pass	3:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:18pm
SRC	Pass	3:18pm
DET	Pass	3:18pm
BAR	Pass	3:18pm
BT	Pass	3:18pm

Blank Tests

Test	Status	Time
AIR	Pass	3:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:19pm

CRC Tests

Test	Status	Time
COMP	Pass	3:19pm
CAL	Pass	3:19pm

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location UNC - Greensboro
Instrument Serial No. 008604 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY UNC-G POLICE DEPT 400

Serial Number: 008604
Test Date: 02/26/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:23am
AIR BLK	.00	11:24am
ACCY CHK	.08	11:25am
AIR BLK	.00	11:26am
SUB TEST	.00	11:27am
AIR BLK	.00	11:28am
SUB TEST	.00	11:29am
AIR BLK	.00	11:30am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY UNC-G POLICE DEPT 400

Serial Number: 008604 Test Record Number: 1607
Test Date: 02/26/2018 Test Time: 11:34am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:34am
FLO	Pass	11:34am
FC	Pass	11:35am

Temperature Tests

Test	Status	Time
FC1	Pass	11:35am
SRC	Pass	11:35am
DET	Pass	11:35am
BAR	Pass	11:35am
BT	Pass	11:35am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am


Printer Tests

Test	Status	Time
PRNT	Pass	11:35am

CRC Tests

Test	Status	Time
COMP	Pass	11:35am
CAL	Pass	11:35am

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Harnett Instrument Location BAT Mobile Unit 12
Instrument Serial No. 008600 Dunn PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of February, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

660
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY BAT MOBILE UNIT 12 420

Serial Number: 008600

Test Date: 02/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, JASON R

Permit Number: 19145E

Effective:

03/01/2016-03/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	10:14pm
AIR BLK	.00	10:15pm
ACCY CHK	.07	10:16pm
AIR BLK	.00	10:17pm
SUB TEST	.00	10:17pm
AIR BLK	.00	10:18pm
SUB TEST	.00	10:20pm
AIR BLK	.00	10:21pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY BAT MOBILE UNIT 12 420

Serial Number: 008600 Test Record Number: 1793
Test Date: 02/16/2018 Test Time: 10:23pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:23pm
FLO	Pass	10:23pm
FC	Pass	10:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:23pm
SRC	Pass	10:23pm
DET	Pass	10:23pm
BAR	Pass	10:23pm
BT	Pass	10:23pm

Blank Tests

Test	Status	Time
AIR	Pass	10:24pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:24pm

CRC Tests

Test	Status	Time
COMP	Pass	10:24pm
CAL	Pass	10:24pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County HARNETT Co. Instrument Location HARNETT Co. Det. Center
Instrument Serial No. 008730 L. Hingston, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of FEBRUARY, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008730

Test Date: 02/28/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: **KEESLER, GRAYHAM C**

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	9:24am
AIR BLK	.00	9:25am
ACCY CHK	.07	9:26am
AIR BLK	.00	9:27am
SUB TEST	.00	9:27am
AIR BLK	.00	9:28am
SUB TEST	.00	9:30am
AIR BLK	.00	9:31am

Reported As: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008730 Test Record Number: 2922
Test Date: 02/28/2018 Test Time: 9:32am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:33am
FLO	Pass	9:33am
FC	Pass	9:33am

Temperature Tests

Test	Status	Time
FC1	Pass	9:33am
SRC	Pass	9:33am
DET	Pass	9:33am
BAR	Pass	9:33am
BT	Pass	9:33am

Blank Tests

Test	Status	Time
AIR	Pass	9:33am


Printer Tests

Test	Status	Time
PRNT	Pass	9:33am

CRC Tests

Test	Status	Time
COMP	Pass	9:34am
CAL	Pass	9:34am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County HARNETT Co. Instrument Location HARNETT Co. Det. Center

Instrument Serial No. 008729 L. Hington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008729

Test Date: 02/28/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302


Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	9:20am
AIR BLK	.00	9:21am
ACCY CHK	.08	9:21am
AIR BLK	.00	9:22am
SUB TEST	.00	9:23am
AIR BLK	.00	9:25am
SUB TEST	.00	9:26am
AIR BLK	.00	9:27am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008729 Test Record Number: 2224
Test Date: 02/28/2018 Test Time: 9:28am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:28am
FLO	Pass	9:28am
FC	Pass	9:28am

Temperature Tests

Test	Status	Time
FC1	Pass	9:28am
SRC	Pass	9:28am
DET	Pass	9:28am
BAR	Pass	9:28am
BT	Pass	9:28am

Blank Tests

Test	Status	Time
AIR	Pass	9:29am


Printer Tests

Test	Status	Time
PRNT	Pass	9:29am

CRC Tests

Test	Status	Time
COMP	Pass	9:29am
CAL	Pass	9:29am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County HARNETT Co. Instrument Location DUNN Police Dept.
Instrument Serial No. 008644 DUNN, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of FEBRUARY, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: 008644

Test Date: 02/28/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:09am
AIR BLK	.00	10:10am
ACCY CHK	.08	10:10am
AIR BLK	.00	10:12am
SUB TEST	.00	10:12am
AIR BLK	.00	10:13am
SUB TEST	.00	10:15am
AIR BLK	.00	10:16am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: 008644 Test Record Number: 1286
Test Date: 02/28/2018 Test Time: 10:17am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:17am
FLO	Pass	10:17am
FC	Pass	10:17am

Temperature Tests

Test	Status	Time
FC1	Pass	10:17am
SRC	Pass	10:17am
DET	Pass	10:17am
BAR	Pass	10:17am
BT	Pass	10:17am

Blank Tests

Test	Status	Time
AIR	Pass	10:18am


Printer Tests

Test	Status	Time
PRNT	Pass	10:18am

CRC Tests

Test	Status	Time
COMP	Pass	10:18am
CAL	Pass	10:18am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson Instrument Location Henderson Co Detention
Instrument Serial No. 008822 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.




Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DETENTION 440

Serial Number: 008822

Test Date: 02/01/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	3:08pm
AIR BLK	.00	3:09pm
ACCY CHK	.08	3:10pm
AIR BLK	.00	3:11pm
SUB TEST	.00	3:11pm
AIR BLK	.00	3:12pm
SUB TEST	.00	3:14pm
AIR BLK	.00	3:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DETENTION 440

Serial Number: 008822 Test Record Number: 2178
Test Date: 02/01/2018 Test Time: 3:16pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:16pm
FLO	Pass	3:16pm
FC	Pass	3:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:16pm
SRC	Pass	3:16pm
DET	Pass	3:16pm
BAR	Pass	3:16pm
BT	Pass	3:16pm

Blank Tests

Test	Status	Time
AIR	Pass	3:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:17pm

CRC Tests

Test	Status	Time
COMP	Pass	3:17pm
CAL	Pass	3:17pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Hertford Instrument Location Murfreesboro P.D.

Instrument Serial No. 008906 115 E. Broad St., Murfreesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: 008906

Test Date: 02/22/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	11:46am
AIR BLK	.00	11:47am
ACCY CHK	.08	11:47am
AIR BLK	.00	11:49am
SUB TEST	.00	11:49am
AIR BLK	.00	11:50am
SUB TEST	.00	11:52am
AIR BLK	.00	11:53am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: 008906 Test Record Number: 654
Test Date: 02/22/2018 Test Time: 11:54am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:55am
FLO	Pass	11:55am
FC	Pass	11:55am

Temperature Tests

Test	Status	Time
FC1	Pass	11:55am
SRC	Pass	11:55am
DET	Pass	11:55am
BAR	Pass	11:55am
BT	Pass	11:55am

Blank Tests

Test	Status	Time
AIR	Pass	11:56am

Printer Tests

Test	Status	Time
PRNT	Pass	11:56am

CRC Tests

Test	Status	Time
COMP	Pass	11:56am
CAL	Pass	11:56am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Iredell Instrument Location Iredell County SD
Instrument Serial No. 008927 201 E Water St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: 008927

Test Date: 02/05/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	3:52pm
AIR BLK	.00	3:52pm
ACCY CHK	.07	3:53pm
AIR BLK	.00	3:54pm
SUB TEST	.00	3:54pm
AIR BLK	.00	3:55pm
SUB TEST	.00	3:57pm
AIR BLK	.00	3:58pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: 008927 Test Record Number: 525
Test Date: 02/05/2018 Test Time: 3:58pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:59pm
FLO	Pass	3:59pm
FC	Pass	3:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:59pm
SRC	Pass	3:59pm
DET	Pass	3:59pm
BAR	Pass	3:59pm
BT	Pass	3:59pm

Blank Tests

Test	Status	Time
AIR	Pass	4:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:00pm

CRC Tests

Test	Status	Time
COMP	Pass	4:00pm
CAL	Pass	4:00pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Johnson Co. Instrument Location Selma Police Dept
Instrument Serial No. 008595 Selma, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26th day of FEBRUARY, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

634
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY SELMA PD 500

Serial Number: 008595

Test Date: 02/26/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:52am
ACCY CHK	.07	10:52am
AIR BLK	.00	10:54am
SUB TEST	.00	10:54am
AIR BLK	.00	10:55am
SUB TEST	.00	10:57am
AIR BLK	.00	10:58am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY SELMA PD 500

Serial Number: 008595 Test Record Number: 1225
Test Date: 02/26/2018 Test Time: 11:00am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:01am
FLO	Pass	11:01am
FC	Pass	11:01am

Temperature Tests

Test	Status	Time
FC1	Pass	11:01am
SRC	Pass	11:01am
DET	Pass	11:01am
BAR	Pass	11:01am
BT	Pass	11:01am

Blank Tests

Test	Status	Time
AIR	Pass	11:02am

Printer Tests

Test	Status	Time
PRNT	Pass	11:02am

CRC Tests

Test	Status	Time
COMP	Pass	11:02am
CAL	Pass	11:02am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Johnson Co. Instrument Location Benson Police Dept.
Instrument Serial No. 008885 Benson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY BENSON POLICE DEPT.
500

Serial Number: 008885
Test Date: 02/28/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501


Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	10:49am
AIR BLK	.00	10:50am
ACCY CHK	.07	10:50am
AIR BLK	.00	10:52am
SUB TEST	.00	10:52am
AIR BLK	.00	10:53am
SUB TEST	.00	10:55am
AIR BLK	.00	10:56am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY BENSON POLICE DEPT. 500

Serial Number: 008885 Test Record Number: 512
Test Date: 02/28/2018 Test Time: 10:56am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:56am
FLO	Pass	10:56am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:57am


Printer Tests

Test	Status	Time
PRNT	Pass	10:57am

CRC Tests

Test	Status	Time
COMP	Pass	10:58am
CAL	Pass	10:58am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lee Co. Instrument Location Lee Co. Jail
Instrument Serial No. 008863 SANFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LEE COUNTY LEE LEC 520

Serial Number: 008863

Test Date: 02/02/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	4:05pm
AIR BLK	.00	4:05pm
ACCY CHK	.07	4:06pm
AIR BLK	.00	4:07pm
SUB TEST	.00	4:07pm
AIR BLK	.00	4:08pm
SUB TEST	.00	4:10pm
AIR BLK	.00	4:10pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY LEE LEC 520

Serial Number: 008863 Test Record Number: 486
Test Date: 02/02/2018 Test Time: 3:59pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:59pm
FLO	Pass	3:59pm
FC	Pass	3:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:59pm
SRC	Pass	3:59pm
DET	Pass	3:59pm
BAR	Pass	3:59pm
BT	Pass	3:59pm

Blank Tests

Test	Status	Time
AIR	Pass	4:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:00pm

CRC Tests

Test	Status	Time
COMP	Pass	4:00pm
CAL	Pass	4:00pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lee Co. Instrument Location Lee Co. Jail
Instrument Serial No. 008645 SANFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LEE COUNTY LEE CO. LEC. 520

Serial Number: 008645

Test Date: 02/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test g/210L Time

DIAG Pass 3:57pm

AIR BLK .00 3:57pm

ACCY CHK .07 3:58pm

AIR BLK .00 3:59pm

SUB TEST .00 4:00pm

AIR BLK .00 4:00pm

SUB TEST .00 4:02pm

AIR BLK .00 4:03pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY LEE CO. LEC. 520

Serial Number: 008645 Test Record Number: 1812
Test Date: 02/15/2018 Test Time: 4:03pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:04pm
FLO	Pass	4:04pm
FC	Pass	4:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:04pm
SRC	Pass	4:04pm
DET	Pass	4:04pm
BAR	Pass	4:04pm
BT	Pass	4:04pm

Blank Tests

Test	Status	Time
AIR	Pass	4:04pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:05pm

CRC Tests

Test	Status	Time
COMP	Pass	4:05pm
CAL	Pass	4:05pm

Preventive Maintenance
Status: *Pass*


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Lee Co. Instrument Location Sanford Police Dept.
Instrument Serial No. 008867 Sanford, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.




Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LEE COUNTY SANFORD POLICE DEPT 520

Serial Number: 008867

Test Date: 02/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	7:56am
AIR BLK	.00	7:57am
ACCY CHK	.07	7:57am
AIR BLK	.00	7:58am
SUB TEST	.00	7:59am
AIR BLK	.00	7:59am
SUB TEST	.00	8:01am
AIR BLK	.00	8:02am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY SANFORD POLICE DEPT 520

Serial Number: 008867 Test Record Number: 1034
Test Date: 02/15/2018 Test Time: 8:03am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	8:04am
FLO	Pass	8:04am
FC	Pass	8:04am

Temperature Tests

Test	Status	Time
FC1	Pass	8:04am
SRC	Pass	8:04am
DET	Pass	8:04am
BAR	Pass	8:04am
BT	Pass	8:04am

Blank Tests

Test	Status	Time
AIR	Pass	8:04am


Printer Tests

Test	Status	Time
PRNT	Pass	8:04am

CRC Tests

Test	Status	Time
COMP	Pass	8:05am
CAL	Pass	8:05am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lincoln Instrument Location Lincoln County Courthouse
Instrument Serial No. 008823 #1 Courthouse Square, Lincolnton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Holt 650
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008823

Test Date: 02/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG722408

Exp Date: 08/12/2019


Test	g/210L	Time
DIAG	Pass	1:10pm
AIR BLK	.00	1:11pm
ACCY CHK	.08	1:11pm
AIR BLK	.00	1:12pm
SUB TEST	.00	1:13pm
AIR BLK	.00	1:14pm
SUB TEST	.00	1:15pm
AIR BLK	.00	1:16pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008823 Test Record Number: 1390

Test Date: 02/09/2018 Test Time: 1:17pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:18pm
FLO	Pass	1:18pm
FC	Pass	1:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:18pm
SRC	Pass	1:18pm
DET	Pass	1:18pm
BAR	Pass	1:18pm
BT	Pass	1:18pm

Blank Tests

Test	Status	Time
AIR	Pass	1:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:19pm

CRC Tests

Test	Status	Time
COMP	Pass	1:19pm
CAL	Pass	1:19pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County Martin Instrument Location Martin Co S.O.

Instrument Serial No. 008912 305 E. Main St., Williamston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008912

Test Date: 02/20/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:59am
AIR BLK	.00	10:59am
ACCY CHK	.08	11:00am
AIR BLK	.00	11:01am
SUB TEST	.00	11:02am
AIR BLK	.00	11:03am
SUB TEST	.00	11:05am
AIR BLK	.00	11:05am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008912 Test Record Number: 1340
Test Date: 02/20/2018 Test Time: 11:08am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:08am
FLO	Pass	11:08am
FC	Pass	11:08am

Temperature Tests

Test	Status	Time
FC1	Pass	11:09am
SRC	Pass	11:09am
DET	Pass	11:09am
BAR	Pass	11:09am
BT	Pass	11:09am

Blank Tests

Test	Status	Time
AIR	Pass	11:09am

Printer Tests

Test	Status	Time
PRNT	Pass	11:09am

CRC Tests

Test	Status	Time
COMP	Pass	11:09am
CAL	Pass	11:09am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Montgomery Co. Instrument Location Montgomery Co. Jail
Instrument Serial No. 008709 TRC4, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MONTGOMERY COUNTY MONTGOMERY CO. JAIL
610

Serial Number: 008709

Test Date: 02/14/2018

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/10/2019

Test	g/210L	Time
DIAG	Pass	6:16pm
AIR BLK	.00	6:17pm
ACCY CHK	.08	6:18pm
AIR BLK	.00	6:19pm
SUB TEST	.00	6:20pm
AIR BLK	.00	6:21pm
SUB TEST	.00	6:22pm
AIR BLK	.00	6:23pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MONTGOMERY COUNTY MONTGOMERY CO. JAIL 610

Serial Number: 008709 Test Record Number: 1036
Test Date: 02/14/2018 Test Time: 6:24pm EST
System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	6:24pm
FLO	Pass	6:24pm
FC	Pass	6:24pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:24pm
SRC	Pass	6:24pm
DET	Pass	6:24pm
BAR	Pass	6:24pm
BT	Pass	6:24pm

Blank Tests

Test	Status	Time
AIR	Pass	6:25pm


Printer Tests

Test	Status	Time
PRNT	Pass	6:25pm

CRC Tests

Test	Status	Time
COMP	Pass	6:25pm
CAL	Pass	6:25pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Montgomery Co. Instrument Location Montgomery Co. Jail
Instrument Serial No. 008657 Troy, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

634
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MONTGOMERY COUNTY MONTGOMERY CO. JAIL
610

Serial Number: 008657
Test Date: 02/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	6:15pm
AIR BLK	.00	6:16pm
ACCY CHK	.07	6:17pm
AIR BLK	.00	6:18pm
SUB TEST	.00	6:19pm
AIR BLK	.00	6:19pm
SUB TEST	.00	6:21pm
AIR BLK	.00	6:22pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MONTGOMERY COUNTY MONTGOMERY CO. JAIL 610

Serial Number: 008657 Test Record Number: 1507
Test Date: 02/14/2018 Test Time: 6:23pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:23pm
FLO	Pass	6:23pm
FC	Pass	6:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:23pm
SRC	Pass	6:23pm
DET	Pass	6:23pm
BAR	Pass	6:23pm
BT	Pass	6:23pm

Blank Tests

Test	Status	Time
AIR	Pass	6:24pm


Printer Tests

Test	Status	Time
PRNT	Pass	6:24pm

CRC Tests

Test	Status	Time
COMP	Pass	6:24pm
CAL	Pass	6:24pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County MOORE Co. Instrument Location Southern Pines P.D.
Instrument Serial No. 008720 Southern Pines, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MOORE COUNTY SOUTHERN PINES PD. 620

Serial Number: 008720

Test Date: 02/07/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test g/210L Time

DIAG	Pass	7:39pm
AIR BLK	.00	7:39pm
ACCY CHK	.07	7:40pm
AIR BLK	.00	7:41pm
SUB TEST	.00	7:41pm
AIR BLK	.00	7:42pm
SUB TEST	.00	7:44pm
AIR BLK	.00	7:45pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY SOUTHERN PINES PD. 620

Serial Number: 008720 Test Record Number: 1014
Test Date: 02/07/2018 Test Time: 7:47pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	7:47pm
FLO	Pass	7:47pm
FC	Pass	7:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:47pm
SRC	Pass	7:47pm
DET	Pass	7:47pm
BAR	Pass	7:47pm
BT	Pass	7:47pm

Blank Tests

Test	Status	Time
AIR	Pass	7:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	7:48pm

CRC Tests

Test	Status	Time
COMP	Pass	7:48pm
CAL	Pass	7:48pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ORANGE

Instrument Location Hillsborough PD

Instrument Serial No. 008873

127 N. Chilton ST

Hillsborough, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Aldes Braxton
Signature of Certifying Official

062
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

~~Intox EC/IR II: Subject Test~~

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008873

Test Date: 02/17/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

~~Agency: BHHS~~


Test Type: Breath Test

Lot Number: AG716202


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	10:17am
AIR BLK	.00	10:18am
ACCY CHK	.08	10:19am
AIR BLK	.00	10:20am
SUB TEST	.00	10:20am
AIR BLK	.00	10:21am
SUB TEST	.00	10:22am
AIR BLK	.00	10:23am

Reported AC: .00 g/210L


Signature of Chemical Analyst

~~Signature of Analyst~~
Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008873 Test Record Number: 1490
Test Date: 02/17/2018 Test Time: 10:26am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:26am
FLO	Pass	10:26am
FC	Pass	10:26am

Temperature Tests

Test	Status	Time
FC1	Pass	10:26am
SRC	Pass	10:26am
DET	Pass	10:26am
BAR	Pass	10:26am
BT	Pass	10:26am

Blank Tests

Test	Status	Time
AIR	Pass	10:27am

Printer Tests


Test	Status	Time
PRNT	Pass	10:27am

CRC Tests

Test	Status	Time
COMP	Pass	10:27am
CAL	Pass	10:27am

Preventive Maintenance

Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Pasquotank Instrument Location Pasquotank Co. Public Safety
Instrument Serial No. 008950 Bldg., 200 E. Colonial St., Elizabeth City
N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Kuehn
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PASQUOTANK COUNTY PUBLIC SAFETY BLDG
690

Serial Number: 008950
Test Date: 02/21/2018

Citation Number: M00000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018


Test	g/210L	Time
DIAG	Pass	10:31am
AIR BLK	.00	10:32am
ACCY CHK	.08	10:33am
AIR BLK	.00	10:34am
SUB TEST	.00	10:34am
AIR BLK	.00	10:35am
SUB TEST	.00	10:37am
AIR BLK	.00	10:38am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY PUBLIC SAFETY BLDG 690

Serial Number: 008950 Test Record Number: 1605
Test Date: 02/21/2018 Test Time: 10:38am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:39am
FLO	Pass	10:39am
FC	Pass	10:39am

Temperature Tests

Test	Status	Time
FC1	Pass	10:39am
SRC	Pass	10:39am
DET	Pass	10:39am
BAR	Pass	10:39am
BT	Pass	10:39am

Blank Tests

Test	Status	Time
AIR	Pass	10:40am


Printer Tests

Test	Status	Time
PRNT	Pass	10:40am

CRC Tests

Test	Status	Time
COMP	Pass	10:40am
CAL	Pass	10:40am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pasquotank Instrument Location Elizabeth City P.D.
Instrument Serial No. 008941 305 E. MAIN ST., Elizabeth City N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21ST day of FEBRUARY, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PASQUOTANK COUNTY ELIZABETH CITY 690

Serial Number: 008941

Test Date: 02/21/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:56am
AIR BLK	.00	10:57am
ACCY CHK	.08	10:58am
AIR BLK	.00	10:59am
SUB TEST	.00	10:59am
AIR BLK	.00	11:00am
SUB TEST	.00	11:02am
AIR BLK	.00	11:03am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY ELIZABETH CITY 690

Serial Number: 008941 Test Record Number: 1270
Test Date: 02/21/2018 Test Time: 11:10am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:11am
FLO	Pass	11:11am
FC	Pass	11:11am

Temperature Tests

Test	Status	Time
FC1	Pass	11:11am
SRC	Pass	11:11am
DET	Pass	11:11am
BAR	Pass	11:11am
BT	Pass	11:11am

Blank Tests

Test	Status	Time
AIR	Pass	11:12am


Printer Tests

Test	Status	Time
PRNT	Pass	11:12am

CRC Tests

Test	Status	Time
COMP	Pass	11:12am
CAL	Pass	11:12am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County PERSON Instrument Location PERSON CO. LEC
Instrument Serial No. 008880 120 COURT ST. ROXBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 06 day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008880
Test Date: 02/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702302

Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	3:48pm
AIR BLK	.00	3:49pm
ACCY CHK	.08	3:50pm
AIR BLK	.00	3:51pm
SUB TEST	.00	3:51pm
AIR BLK	.00	3:52pm
SUB TEST	.00	3:54pm
AIR BLK	.00	3:54pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008880 Test Record Number: 1259
Test Date: 02/06/2018 Test Time: 3:55pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:56pm
FLO	Pass	3:56pm
FC	Pass	3:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:56pm
SRC	Pass	3:56pm
DET	Pass	3:56pm
BAR	Pass	3:56pm
BT	Pass	3:56pm

Blank Tests

Test	Status	Time
AIR	Pass	3:56pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:57pm

CRC Tests

Test	Status	Time
COMP	Pass	3:57pm
CAL	Pass	3:57pm

**Preventive Maintenance
Status: Pass**


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County PERSON

Instrument Location PERSON CO. LEC

Instrument Serial No. 008693

120 COURT ST. ROXBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 06 day of FEBRUARY, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008693
Test Date: 02/06/2018

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702302


Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	3:47pm
AIR BLK	.00	3:48pm
ACCY CHK	.07	3:48pm
AIR BLK	.00	3:49pm
SUB TEST	.00	3:50pm
AIR BLK	.00	3:51pm
SUB TEST	.00	3:52pm
AIR BLK	.00	3:53pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008693 Test Record Number: 1427
Test Date: 02/06/2018 Test Time: 3:54pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:55pm
FLO	Pass	3:55pm
FC	Pass	3:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:55pm
SRC	Pass	3:55pm
DET	Pass	3:55pm
BAR	Pass	3:55pm
BT	Pass	3:55pm

Blank Tests

Test	Status	Time
AIR	Pass	3:56pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:56pm

CRC Tests

Test	Status	Time
COMP	Pass	3:56pm
CAL	Pass	3:56pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Instrument Location Bat Mobile Unit 12
Instrument Serial No. 008788 Randolph CO SD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donya B Skimin
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY BAT MOBILE UNIT 12 750

Serial Number: 008788
Test Date: 02/24/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911.

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	11:21pm
AIR BLK	.00	11:22pm
ACCY CHK	.07	11:23pm
AIR BLK	.00	11:24pm
SUB TEST	.00	11:24pm
AIR BLK	.00	11:25pm
SUB TEST	.00	11:27pm
AIR BLK	.00	11:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY BAT MOBILE UNIT 12 750

Serial Number: 008788

Test Record Number: 1326

Test Date: 02/24/2018

Test Time: 11:30pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:30pm
FLO	Pass	11:30pm
FC	Pass	11:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:31pm
SRC	Pass	11:31pm
DET	Pass	11:31pm
BAR	Pass	11:31pm
BT	Pass	11:31pm

Blank Tests

Test	Status	Time
AIR	Pass	11:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:31pm

CRC Tests

Test	Status	Time
COMP	Pass	11:31pm
CAL	Pass	11:31pm

Preventive Maintenance

Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Instrument Location Bat Mobile Unit 12
Instrument Serial No. 008698 Randolph CO 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY BAT MOBILE UNIT 12 750

Serial Number: 008698

Test Date: 02/25/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	12:21am
AIR BLK	.00	12:22am
ACCY CHK	.07	12:22am
AIR BLK	.00	12:23am
SUB TEST	.00	12:25am
AIR BLK	.00	12:26am
SUB TEST	.00	12:27am
AIR BLK	.00	12:28am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY BAT MOBILE UNIT 12 750

Serial Number: 008698 Test Record Number: 1328
Test Date: 02/25/2018 Test Time: 12:30am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:30am
FLO	Pass	12:30am
FC	Pass	12:30am

Temperature Tests

Test	Status	Time
FC1	Pass	12:30am
SRC	Pass	12:30am
DET	Pass	12:30am
BAR	Pass	12:30am
BT	Pass	12:30am

Blank Tests

Test	Status	Time
AIR	Pass	12:31am


Printer Tests

Test	Status	Time
PRNT	Pass	12:31am

CRC Tests

Test	Status	Time
COMP	Pass	12:31am
CAL	Pass	12:31am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Instrument Location Ardenale
Instrument Serial No. 008791 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791

Test Date: 02/27/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

04/01/2017-04/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:13am
AIR BLK	.00	11:13am
ACCY CHK	.08	11:14am
AIR BLK	.00	11:15am
SUB TEST	.00	11:16am
AIR BLK	.00	11:17am
SUB TEST	.00	11:18am
AIR BLK	.00	11:19am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791 Test Record Number: 1241
Test Date: 02/27/2018 Test Time: 11:22am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:22am
FLO	Pass	11:22am
FC	Pass	11:23am

Temperature Tests

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

Blank Tests

Test	Status	Time
AIR	Pass	11:23am


Printer Tests

Test	Status	Time
PRNT	Pass	11:23am

CRC Tests

Test	Status	Time
COMP	Pass	11:24am
CAL	Pass	11:24am

Preventive Maintenance
Status: Pass


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rockingham Instrument Location Madison Police
Instrument Serial No. 008802 Department ✓

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: 008802

Test Date: 02/08/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	4:13pm
AIR BLK	.00	4:14pm
ACCY CHK	.08	4:15pm
AIR BLK	.00	4:16pm
SUB TEST	.00	4:17pm
AIR BLK	.00	4:18pm
SUB TEST	.00	4:19pm
AIR BLK	.00	4:20pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: 008802 Test Record Number: 760

Test Date: 02/08/2018 Test Time: 4:21pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	4:22pm
FLO	Pass	4:22pm
FC	Pass	4:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:22pm
SRC	Pass	4:22pm
DET	Pass	4:22pm
BAR	Pass	4:22pm
BT	Pass	4:22pm

Blank Tests

Test	Status	Time
AIR	Pass	4:23pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:23pm

CRC Tests

Test	Status	Time
COMP	Pass	4:23pm
CAL	Pass	4:23pm

**Preventive Maintenance
Status: Pass**


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rutherford Instrument Location Rutherford County SD
Instrument Serial No. 008914 400 N. Washington St., Rutherford

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RUTHERFORD COUNTY RUTHERFORD COUNTY SD
800

Serial Number: 008914
Test Date: 02/20/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test g/210L Time

DIAG	Pass	12:39pm
AIR BLK	.00	12:40pm
ACCY CHK	.07	12:41pm
AIR BLK	.00	12:42pm
SUB TEST	.00	12:42pm
AIR BLK	.00	12:43pm
SUB TEST	.00	12:45pm
AIR BLK	.00	12:46pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

RUTHERFORD COUNTY RUTHERFORD COUNTY SD 800

Serial Number: 008914 Test Record Number: 1956

Test Date: 02/20/2018 Test Time: 12:48pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:49pm
FLO	Pass	12:49pm
FC	Pass	12:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:49pm
SRC	Pass	12:49pm
DET	Pass	12:49pm
BAR	Pass	12:49pm
BT	Pass	12:49pm

Blank Tests

Test	Status	Time
AIR	Pass	12:49pm

Printer Tests


Test	Status	Time
PRNT	Pass	12:50pm

CRC Tests

Test	Status	Time
COMP	Pass	12:50pm
CAL	Pass	12:50pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Scotland Co. Instrument Location Laurinburg Police Dept.
Instrument Serial No. 008834 Laurinburg P.D.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SCOTLAND COUNTY LAURINBURG PD. 820

Serial Number: 008834

Test Date: 02/07/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	6:36pm
AIR BLK	.00	6:37pm
ACCY CHK	.07	6:37pm
AIR BLK	.00	6:38pm
SUB TEST	.00	6:39pm
AIR BLK	.00	6:40pm
SUB TEST	.00	6:42pm
AIR BLK	.00	6:43pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

SCOTLAND COUNTY LAURINBURG PD. 820

Serial Number: 008834 Test Record Number: 833
Test Date: 02/07/2018 Test Time: 6:43pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:43pm
FLO	Pass	6:43pm
FC	Pass	6:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:43pm
SRC	Pass	6:43pm
DET	Pass	6:43pm
BAR	Pass	6:43pm
BT	Pass	6:43pm

Blank Tests

Test	Status	Time
AIR	Pass	6:44pm


Printer Tests

Test	Status	Time
PRNT	Pass	6:44pm

CRC Tests

Test	Status	Time
COMP	Pass	6:44pm
CAL	Pass	6:44pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Scotland Co. Instrument Location Scotland Co. Jail
Instrument Serial No. 008861 LAURINBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861

Test Date: 02/07/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	6:03pm
AIR BLK	.00	6:04pm
ACCY CHK	.07	6:04pm
AIR BLK	.00	6:05pm
SUB TEST	.00	6:06pm
AIR BLK	.00	6:07pm
SUB TEST	.00	6:08pm
AIR BLK	.00	6:09pm

Reported AC .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861 Test Record Number: 1377
Test Date: 02/07/2018 Test Time: 6:10pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	6:10pm
FLO	Pass	6:10pm
FC	Pass	6:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:10pm
SRC	Pass	6:10pm
DET	Pass	6:10pm
BAR	Pass	6:10pm
BT	Pass	6:10pm

Blank Tests

Test	Status	Time
AIR	Pass	6:11pm


Printer Tests

Test	Status	Time
PRNT	Pass	6:11pm

CRC Tests

Test	Status	Time
COMP	Pass	6:11pm
CAL	Pass	6:11pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location Stokes County Jail
Instrument Serial No. 008596 Danbury, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: 008596

Test Date: 02/08/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401


Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	5:08pm
AIR BLK	.00	5:09pm
ACCY CHK	.08	5:09pm
AIR BLK	.00	5:11pm
SUB TEST	.00	5:11pm
AIR BLK	.00	5:12pm
SUB TEST	.00	5:13pm
AIR BLK	.00	5:14pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: 008596 Test Record Number: 984
Test Date: 02/08/2018 Test Time: 5:15pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	5:16pm
FLO	Pass	5:16pm
FC	Pass	5:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:16pm
SRC	Pass	5:16pm
DET	Pass	5:16pm
BAR	Pass	5:16pm
BT	Pass	5:16pm

Blank Tests

Test	Status	Time
AIR	Pass	5:17pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:17pm

CRC Tests

Test	Status	Time
COMP	Pass	5:17pm
CAL	Pass	5:17pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Stokes Instrument Location King Police
Instrument Serial No. 008610 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of February, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STOKES COUNTY KING PD 840

Serial Number: 008610

Test Date: 02/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401


Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	4:57pm
AIR BLK	.00	4:57pm
ACCY CHK	.07	4:58pm
AIR BLK	.00	4:59pm
SUB TEST	.00	5:00pm
AIR BLK	.00	5:01pm
SUB TEST	.00	5:02pm
AIR BLK	.00	5:03pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY KING PD 840

Serial Number: 008610 Test Record Number: 1922
Test Date: 02/09/2018 Test Time: 5:04pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	5:05pm
FLO	Pass	5:05pm
FC	Pass	5:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:05pm
SRC	Pass	5:05pm
DET	Pass	5:05pm
BAR	Pass	5:05pm
BT	Pass	5:05pm

Blank Tests

Test	Status	Time
AIR	Pass	5:05pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:05pm

CRC Tests

Test	Status	Time
COMP	Pass	5:06pm
CAL	Pass	5:06pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County

Surry

Instrument Location

Elkin Police

Instrument Serial No.

008926

Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of February, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SURRY COUNTY ELKIN PD 850

Serial Number: 008926

Test Date: 02/12/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734102


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	8:37am
AIR BLK	.00	8:38am
ACCY CHK	.08	8:39am
AIR BLK	.00	8:40am
SUB TEST	.00	8:41am
AIR BLK	.00	8:42am
SUB TEST	.00	8:43am
AIR BLK	.00	8:44am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY ELKIN PD 850

Serial Number: 008926 Test Record Number: 786
Test Date: 02/12/2018 Test Time: 8:45am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	8:45am
FLO	Pass	8:45am
FC	Pass	8:45am

Temperature Tests

Test	Status	Time
FC1	Pass	8:45am
SRC	Pass	8:45am
DET	Pass	8:45am
BAR	Pass	8:45am
BT	Pass	8:45am

Blank Tests

Test	Status	Time
AIR	Pass	8:46am


Printer Tests

Test	Status	Time
PRNT	Pass	8:46am

CRC Tests

Test	Status	Time
COMP	Pass	8:46am
CAL	Pass	8:46am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Surry Instrument Location Pilot Mountain
Instrument Serial No. 008938 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SURRY COUNTY PILOT MOUNTAIN PD 850

Serial Number: 008938

Test Date: 02/14/2018

Citation Number: M00000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	3:14pm
AIR BLK	.00	3:15pm
ACCY CHK	.08	3:16pm
AIR BLK	.00	3:17pm
SUB TEST	.00	3:18pm
AIR BLK	.00	3:19pm
SUB TEST	.00	3:22pm
AIR BLK	.00	3:23pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY PILOT MOUNTAIN PD 850

Serial Number: 008938 Test Record Number: 597
Test Date: 02/14/2018 Test Time: 3:24pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:25pm
FLO	Pass	3:25pm
FC	Pass	3:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:25pm
SRC	Pass	3:25pm
DET	Pass	3:25pm
BAR	Pass	3:25pm
BT	Pass	3:25pm

Blank Tests

Test	Status	Time
AIR	Pass	3:26pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:26pm

CRC Tests

Test	Status	Time
COMP	Pass	3:26pm
CAL	Pass	3:26pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Swain Instrument Location Cherokee Tribal Det.
Instrument Serial No. 008782 Cherokee, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SWAIN COUNTY CHEROKEE DETENTION 860

Serial Number: 008782

Test Date: 02/22/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621501


Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	11:56am
AIR BLK	.00	11:57am
ACCY CHK	.08	11:58am
AIR BLK	.00	11:59am
SUB TEST	.00	11:59am
AIR BLK	.00	12:00pm
SUB TEST	.00	12:02pm
AIR BLK	.00	12:02pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SWAIN COUNTY CHEROKEE DETENTION 860

Serial Number: 008782 Test Record Number: 1029

Test Date: 02/22/2018 Test Time: 12:04pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:04pm
FLO	Pass	12:04pm
FC	Pass	12:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:04pm
SRC	Pass	12:04pm
DET	Pass	12:04pm
BAR	Pass	12:04pm
BT	Pass	12:04pm

Blank Tests

Test	Status	Time
AIR	Pass	12:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:05pm

CRC Tests

Test	Status	Time
COMP	Pass	12:05pm
CAL	Pass	12:05pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County VANCE Instrument Location VANCE CO SHERIFF'S OFFICE

Instrument Serial No. 008937 156 CHURCH ST. HENDERSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 05 day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937

Test Date: 02/05/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702402

Exp Date: 01/24/2019


Test	g/210L	Time
DIAG	Pass	11:54am
AIR BLK	.00	11:55am
ACCY CHK	.08	11:56am
AIR BLK	.00	11:57am
SUB TEST	.00	11:57am
AIR BLK	.00	11:58am
SUB TEST	.00	12:00pm
AIR BLK	.00	12:01pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937 Test Record Number: 2426
Test Date: 02/05/2018 Test Time: 12:02pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:03pm
FLO	Pass	12:03pm
FC	Pass	12:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:03pm
SRC	Pass	12:03pm
DET	Pass	12:03pm
BAR	Pass	12:03pm
BT	Pass	12:03pm

Blank Tests

Test	Status	Time
AIR	Pass	12:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:04pm

CRC Tests

Test	Status	Time
COMP	Pass	12:04pm
CAL	Pass	12:04pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County VANCE

Instrument Location VANCE CO. SHERIFF'S OFFICE


Instrument Serial No. 008870

156 CHURCH ST. HENDERSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 05 day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



B. J. Smith

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870

Test Date: 02/05/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019


Test	g/210L	Time
DIAG	Pass	11:51am
AIR BLK	.00	11:52am
ACCY CHK	.07	11:52am
AIR BLK	.00	11:54am
SUB TEST	.00	11:55am
AIR BLK	.00	11:56am
SUB TEST	.00	11:57am
AIR BLK	.00	11:58am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870 Test Record Number: 2167
Test Date: 02/05/2018 Test Time: 11:59am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:59am
FLO	Pass	11:59am
FC	Pass	11:59am

Temperature Tests

Test	Status	Time
FC1	Pass	12:00pm
SRC	Pass	12:00pm
DET	Pass	12:00pm
BAR	Pass	12:00pm
BT	Pass	12:00pm

Blank Tests

Test	Status	Time
AIR	Pass	12:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:00pm

CRC Tests

Test	Status	Time
COMP	Pass	12:00pm
CAL	Pass	12:00pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location CARY PD

Instrument Serial No. 008587 120 Wilkinson Ave
CARY, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CARY PD 910

Serial Number: 008587

Test Date: 02/01/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:53pm
AIR BLK	.00	1:54pm
ACCY CHK	.08	1:55pm
AIR BLK	.00	1:56pm
SUB TEST	.00	1:57pm
AIR BLK	.00	1:58pm
SUB TEST	.00	1:59pm
AIR BLK	.00	2:00pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CARY PD 910

Serial Number: 008587 Test Record Number: 3676
Test Date: 02/01/2018 Test Time: 2:01pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:02pm
FLO	Pass	2:02pm
FC	Pass	2:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:02pm
SRC	Pass	2:02pm
DET	Pass	2:02pm
BAR	Pass	2:02pm
BT	Pass	2:02pm

Blank Tests

Test	Status	Time
AIR	Pass	2:02pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:02pm

CRC Tests

Test	Status	Time
COMP	Pass	2:03pm
CAL	Pass	2:03pm

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Apex PD Station 4
Instrument Serial No. 008621 1615 E. Williams St. Apex, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of FEBRUARY, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY APEX PD

Serial Number: 008621

Test Date: 02/01/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NNE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	11:30am
AIR BLK	.00	11:31am
ACCY CHK	.07	11:32am
AIR BLK	.00	11:33am
SUB TEST	.00	11:33am
AIR BLK	.00	11:34am
SUB TEST	.00	11:36am
AIR BLK	.00	11:37am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY APEX PD

Serial Number: 008621 Test Record Number: 2387

Test Date: 02/01/2018 Test Time: 11:37am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:37am
FLO	Pass	11:37am
FC	Pass	11:37am

Temperature Tests

Test	Status	Time
FC1	Pass	11:38am
SRC	Pass	11:38am
DET	Pass	11:38am
BAR	Pass	11:38am
BT	Pass	11:38am

Blank Tests

Test	Status	Time
AIR	Pass	11:38am


Printer Tests

Test	Status	Time
PRNT	Pass	11:38am

CRC Tests

Test	Status	Time
COMP	Pass	11:38am
CAL	Pass	11:38am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE


Instrument Location KNIGHTDALE P. S.

Instrument Serial No. 008838 979 STEEPLE SQUARE CT.
KNIGHTDALE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 01 day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY KNIGHTDALE PS 910

Serial Number: 008838

Test Date: 02/01/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	3:30pm
AIR BLK	.00	3:31pm
ACCY CHK	.07	3:32pm
AIR BLK	.00	3:33pm
SUB TEST	.00	3:33pm
AIR BLK	.00	3:34pm
SUB TEST	.00	3:36pm
AIR BLK	.00	3:37pm

Reported AC: .00 g/210L

Brian D Smith
Signature of Chemical Analyst

Court CVR

Brian D Smith
Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY KNIGHTDALE PS 910

Serial Number: 008838 Test Record Number: 1621
Test Date: 02/01/2018 Test Time: 3:37pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:38pm
FLO	Pass	3:38pm
FC	Pass	3:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:38pm
SRC	Pass	3:38pm
DET	Pass	3:38pm
BAR	Pass	3:38pm
BT	Pass	3:38pm

Blank Tests

Test	Status	Time
AIR	Pass	3:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:39pm

CRC Tests

Test	Status	Time
COMP	Pass	3:39pm
CAL	Pass	3:39pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE CO DETENTION CENTER

Instrument Serial No. UD8651 3301 HAMMOND RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of FEBRUARY, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



B. D. Smith

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008651
Test Date: 02/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:03pm
AIR BLK	.00	1:04pm
ACCY CHK	.07	1:04pm
AIR BLK	.00	1:05pm
SUB TEST	.00	1:06pm
AIR BLK	.00	1:07pm
SUB TEST	.00	1:08pm
AIR BLK	.00	1:09pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008651 Test Record Number: 1291

Test Date: 02/09/2018 Test Time: 1:12pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:12pm
FLO	Pass	1:12pm
FC	Pass	1:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:12pm
SRC	Pass	1:12pm
DET	Pass	1:12pm
BAR	Pass	1:12pm
BT	Pass	1:12pm

Blank Tests

Test	Status	Time
AIR	Pass	1:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:13pm

CRC Tests

Test	Status	Time
COMP	Pass	1:13pm
CAL	Pass	1:13pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE CO. DETENTION CENTER
Instrument Serial No. 008738 3301 HAMMUND RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008738

Test Date: 02/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	12:32pm
AIR BLK	.00	12:33pm
ACCY CHK	.08	12:33pm
AIR BLK	.00	12:35pm
SUB TEST	.00	12:35pm
AIR BLK	.00	12:36pm
SUB TEST	.00	12:38pm
AIR BLK	.00	12:39pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008738 Test Record Number: 759
Test Date: 02/09/2018 Test Time: 12:40pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:40pm
FLO	Pass	12:40pm
FC	Pass	12:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:40pm
SRC	Pass	12:40pm
DET	Pass	12:40pm
BAR	Pass	12:40pm
BT	Pass	12:40pm

Blank Tests

Test	Status	Time
AIR	Pass	12:41pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:41pm

CRC Tests

Test	Status	Time
COMP	Pass	12:41pm
CAL	Pass	12:41pm

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wake Instrument Location Wake Co Jail

Instrument Serial No. 008577 3301 Hammond Rd
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577

Test Date: 02/19/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:39am
AIR BLK	.00	9:40am
ACCY CHK	.07	9:41am
AIR BLK	.00	9:42am
SUB TEST	.00	9:44am
AIR BLK	.00	9:45am
SUB TEST	.00	9:46am
AIR BLK	.00	9:47am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577 Test Record Number: 3671
Test Date: 02/19/2018 Test Time: 9:51am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:51am
FLO	Pass	9:51am
FC	Pass	9:51am

Temperature Tests

Test	Status	Time
FC1	Pass	9:51am
SRC	Pass	9:51am
DET	Pass	9:51am
BAR	Pass	9:51am
BT	Pass	9:51am

Blank Tests

Test	Status	Time
AIR	Pass	9:52am

Printer Tests

Test	Status	Time
PRNT	Pass	9:52am

CRC Tests

Test	Status	Time
COMP	Pass	9:52am
CAL	Pass	9:52am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE Co JAIL
Instrument Serial No. 008612 3301 Hammond Rd
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of February, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

442
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612

Test Date: 02/19/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716202


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	10:05am
AIR BLK	.00	10:06am
ACCY CHK	.07	10:07am
AIR BLK	.00	10:08am
SUB TEST	.00	10:08am
AIR BLK	.00	10:09am
SUB TEST	.00	10:11am
AIR BLK	.00	10:11am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612 Test Record Number: 3850
Test Date: 02/19/2018 Test Time: 10:16am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:16am
FLO	Pass	10:16am
FC	Pass	10:16am

Temperature Tests

Test	Status	Time
FC1	Pass	10:16am
SRC	Pass	10:16am
DET	Pass	10:16am
BAR	Pass	10:16am
BT	Pass	10:16am

Blank Tests

Test	Status	Time
AIR	Pass	10:17am


Printer Tests

Test	Status	Time
PRNT	Pass	10:17am

CRC Tests

Test	Status	Time
COMP	Pass	10:17am
CAL	Pass	10:17am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County

Wayne

Instrument Location

Seymour Johnson A.F.B.

Instrument Serial No.


008746

1010 Vermont Garrison Rd., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20th day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Tim A. Keel
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786

Test Date: 02/20/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	1:37pm
AIR BLK	.00	1:38pm
ACCY CHK	.07	1:39pm
AIR BLK	.00	1:40pm
SUB TEST	.00	1:40pm
AIR BLK	.00	1:41pm
SUB TEST	.00	1:43pm
AIR BLK	.00	1:44pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786 Test Record Number: 291

Test Date: 02/20/2018 Test Time: 1:45pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:45pm
FLO	Pass	1:45pm
FC	Pass	1:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:45pm
SRC	Pass	1:45pm
DET	Pass	1:45pm
BAR	Pass	1:45pm
BT	Pass	1:45pm

Blank Tests

Test	Status	Time
AIR	Pass	1:46pm

Printer Tests


Test	Status	Time
PRNT	Pass	1:46pm

CRC Tests

Test	Status	Time
COMP	Pass	1:46pm
CAL	Pass	1:46pm

Preventive Maintenance

Status: Pass



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wayne Instrument Location Wayne Co. Detention CTR.

Instrument Serial No. 008649 207 E. CHESTNUT ST., GOLDSBORO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Kessel
Signature of Certifying Official

1047
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649

Test Date: 02/22/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	1:25pm
AIR BLK	.00	1:26pm
ACCY CHK	.08	1:27pm
AIR BLK	.00	1:28pm
SUB TEST	.00	1:28pm
AIR BLK	.00	1:29pm
SUB TEST	.00	1:31pm
AIR BLK	.00	1:32pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649 Test Record Number: 3361

Test Date: 02/22/2018 Test Time: 1:33pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:33pm
FLO	Pass	1:33pm
FC	Pass	1:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

Blank Tests

Test	Status	Time
AIR	Pass	1:34pm

Printer Tests


Test	Status	Time
PRNT	Pass	1:34pm

CRC Tests

Test	Status	Time
COMP	Pass	1:34pm
CAL	Pass	1:34pm

Preventive Maintenance

Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wayne Instrument Location Wayne Co. Detention Ctr.
Instrument Serial No. 008671 207 E. CHESTNUT ST., GOLDSBORO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of FEBRUARY, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Keene
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671

Test Date: 02/22/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019


Test	g/210L	Time
DIAG	Pass	1:28pm
AIR BLK	.00	1:29pm
ACCY CHK	.08	1:30pm
AIR BLK	.00	1:31pm
SUB TEST	.00	1:31pm
AIR BLK	.00	1:32pm
SUB TEST	.00	1:35pm
AIR BLK	.00	1:36pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671 Test Record Number: 4530
Test Date: 02/22/2018 Test Time: 1:39pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:39pm
FLO	Pass	1:39pm
FC	Pass	1:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:40pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:40pm

CRC Tests

Test	Status	Time
COMP	Pass	1:40pm
CAL	Pass	1:40pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Watauga Instrument Location Watauga Co. Jail
Instrument Serial No. 008215 Bone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: 008715

Test Date: 02/27/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG722408

Exp Date: 08/12/2019

Test	g/210L	Time
DIAG	Pass	1:36pm
AIR BLK	.00	1:37pm
ACCY CHK	.08	1:38pm
AIR BLK	.00	1:39pm
SUB TEST	.00	1:40pm
AIR BLK	.00	1:41pm
SUB TEST	.00	1:42pm
AIR BLK	.00	1:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: 008715 Test Record Number: 2079
Test Date: 02/27/2018 Test Time: 1:44pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:44pm
FLO	Pass	1:44pm
FC	Pass	1:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:45pm
SRC	Pass	1:45pm
DET	Pass	1:45pm
BAR	Pass	1:45pm
BT	Pass	1:45pm

Blank Tests

Test	Status	Time
AIR	Pass	1:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:45pm

CRC Tests

Test	Status	Time
COMP	Pass	1:46pm
CAL	Pass	1:46pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wilkes Instrument Location Wilkes County Detention
Instrument Serial No. 008843 Wilkesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843

Test Date: 02/12/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:51am
AIR BLK	.00	9:51am
ACCY CHK	.07	9:52am
AIR BLK	.00	9:53am
SUB TEST	.00	9:54am
AIR BLK	.00	9:55am
SUB TEST	.00	9:56am
AIR BLK	.00	9:57am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843 Test Record Number: 2209
Test Date: 02/12/2018 Test Time: 9:57am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:58am
FLO	Pass	9:58am
FC	Pass	9:58am

Temperature Tests

Test	Status	Time
FC1	Pass	9:58am
SRC	Pass	9:58am
DET	Pass	9:58am
BAR	Pass	9:58am
BT	Pass	9:58am

Blank Tests

Test	Status	Time
AIR	Pass	9:58am


Printer Tests

Test	Status	Time
PRNT	Pass	9:58am

CRC Tests

Test	Status	Time
COMP	Pass	9:59am
CAL	Pass	9:59am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County

Wilkes

Instrument Location

Wilkes County Detention

Instrument Serial No.


008865

Wilkesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of February, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865

Test Date: 02/12/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:47am
AIR BLK	.00	9:48am
ACCY CHK	.07	9:49am
AIR BLK	.00	9:50am
SUB TEST	.00	9:51am
AIR BLK	.00	9:52am
SUB TEST	.00	9:53am
AIR BLK	.00	9:54am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865 Test Record Number: 575
Test Date: 02/12/2018 Test Time: 9:55am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:56am
FLO	Pass	9:56am
FC	Pass	9:56am

Temperature Tests

Test	Status	Time
FC1	Pass	9:56am
SRC	Pass	9:56am
DET	Pass	9:56am
BAR	Pass	9:56am
BT	Pass	9:56am

Blank Tests

Test	Status	Time
AIR	Pass	9:57am


Printer Tests

Test	Status	Time
PRNT	Pass	9:57am

CRC Tests

Test	Status	Time
COMP	Pass	9:57am
CAL	Pass	9:57am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wilson Instrument Location Wilson County Detention Center

Instrument Serial No. 008627 100 E. Green St., Wilson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly A. O.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILSON COUNTY DETENTION CENTER 970

Serial Number: 008627

Test Date: 02/13/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	10:19am
AIR BLK	.00	10:20am
ACCY CHK	.08	10:21am
AIR BLK	.00	10:22am
SUB TEST	.00	10:23am
AIR BLK	.00	10:24am
SUB TEST	.00	10:26am
AIR BLK	.00	10:27am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY DETENTION CENTER 970

Serial Number: 008627 Test Record Number: 2187
Test Date: 02/13/2018 Test Time: 10:28am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:28am
FLO	Pass	10:28am
FC	Pass	10:29am

Temperature Tests

Test	Status	Time
FC1	Pass	10:29am
SRC	Pass	10:29am
DET	Pass	10:29am
BAR	Pass	10:29am
BT	Pass	10:29am

Blank Tests

Test	Status	Time
AIR	Pass	10:29am


Printer Tests

Test	Status	Time
PRNT	Pass	10:29am

CRC Tests

Test	Status	Time
COMP	Pass	10:29am
CAL	Pass	10:29am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wilson Instrument Location Wilson Co. Detention Center

Instrument Serial No. 008652 100 E. Green St., Wilson, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13th day of February, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kenneth O
Signature of Certifying Official

1043
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILSON COUNTY DETENTION CENTER 970

Serial Number: 008652

Test Date: 02/13/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	9:53am
AIR BLK	.00	9:54am
ACCY CHK	.07	9:54am
AIR BLK	.00	9:56am
SUB TEST	.00	9:56am
AIR BLK	.00	9:57am
SUB TEST	.00	9:59am
AIR BLK	.00	10:00am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY DETENTION CENTER 970

Serial Number: 008652 Test Record Number: 3062

Test Date: 02/13/2018 Test Time: 10:13am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:13am
FLO	Pass	10:13am
FC	Pass	10:13am

Temperature Tests

Test	Status	Time
FC1	Pass	10:14am
SRC	Pass	10:14am
DET	Pass	10:14am
BAR	Pass	10:14am
BT	Pass	10:14am

Blank Tests

Test	Status	Time
AIR	Pass	10:14am


Printer Tests

Test	Status	Time
PRNT	Pass	10:14am

CRC Tests

Test	Status	Time
COMP	Pass	10:14am
CAL	Pass	10:14am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County

Yadkin

Instrument Location

Yadkin County Jail
Yadkinville, N.C.

Instrument Serial No.

008944

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944

Test Date: 02/07/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302


Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	5:38pm
AIR BLK	.00	5:38pm
ACCY CHK	.07	5:39pm
AIR BLK	.00	5:40pm
SUB TEST	.00	5:41pm
AIR BLK	.00	5:42pm
SUB TEST	.00	5:43pm
AIR BLK	.00	5:44pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944 Test Record Number: 1473
Test Date: 02/07/2018 Test Time: 5:45pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:46pm
FLO	Pass	5:46pm
FC	Pass	5:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:46pm
SRC	Pass	5:46pm
DET	Pass	5:46pm
BAR	Pass	5:46pm
BT	Pass	5:46pm

Blank Tests

Test	Status	Time
AIR	Pass	5:46pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:46pm

CRC Tests

Test	Status	Time
COMP	Pass	5:47pm
CAL	Pass	5:47pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Yadkin

Instrument Location Yadkin County Jail


Instrument Serial No. 008854

Yadkinville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of February, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008854

Test Date: 02/07/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302

Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	5:30pm
AIR BLK	.00	5:31pm
ACCY CHK	.08	5:31pm
AIR BLK	.00	5:33pm
SUB TEST	.00	5:34pm
AIR BLK	.00	5:35pm
SUB TEST	.00	5:36pm
AIR BLK	.00	5:37pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008854 Test Record Number: 464
Test Date: 02/07/2018 Test Time: 5:39pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:39pm
FLO	Pass	5:39pm
FC	Pass	5:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:39pm
SRC	Pass	5:39pm
DET	Pass	5:39pm
BAR	Pass	5:39pm
BT	Pass	5:39pm

Blank Tests

Test	Status	Time
AIR	Pass	5:40pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:40pm

CRC Tests

Test	Status	Time
COMP	Pass	5:40pm
CAL	Pass	5:40pm

Preventive Maintenance
Status: Pass


Analyst